

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066407.D
 Acq On : 30 Mar 2021 3:55
 Operator : JC/MD
 Sample : M1821-05
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 EL-510-STREET-LEVEL

Quant Time: Mar 30 07:09:42 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	402924	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	615204	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	535922	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.286	152	201967	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	286733	53.53	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 107.06%	
35) Dibromofluoromethane	7.504	113	225309	55.24	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 110.48%	
50) Toluene-d8	10.028	98	836407	54.07	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.14%	
62) 4-Bromofluorobenzene	12.345	95	238654	44.42	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.84%	
Target Compounds						
16) Acetone	3.736	43	102315	35.36	ug/l	97
20) Methylene Chloride	4.449	84	32219	5.13	ug/l	97
43) Isopropyl Acetate	8.094	43	35886	3.30	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066407.D
Acq On : 30 Mar 2021 3:55
Operator : JC/MD
Sample : M1821-05
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
EL-510-STREET-LEVEL

Quant Time: Mar 30 07:09:42 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

