

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
 Data File : VN066409.D
 Acq On : 30 Mar 2021 4:42
 Operator : JC/MD
 Sample : M1832-06
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 49 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IG001

Quant Time: Mar 30 07:10:15 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.584	168	402859	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.512	114	613175	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	531737	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	207311	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.946	65	279519	52.19	ug/l	0.00
Spiked Amount	50.000	Range	78 - 117	Recovery	= 104.38%	
35) Dibromofluoromethane	7.507	113	219033	53.88	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 107.76%	
50) Toluene-d8	10.028	98	832549	54.00	ug/l	0.00
Spiked Amount	50.000	Range	92 - 112	Recovery	= 108.00%	
62) 4-Bromofluorobenzene	12.345	95	237583	44.36	ug/l	0.00
Spiked Amount	50.000	Range	83 - 123	Recovery	= 88.72%	
Target Compounds						Qvalue
16) Acetone	3.736	43	227824	78.74	ug/l	99
18) Methyl Acetate	4.235	43	9488	1.26	ug/l	# 88
20) Methylene Chloride	4.452	84	30741	4.83	ug/l	98
43) Isopropyl Acetate	8.091	43	16920	1.56	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032921\
Data File : VN066409.D
Acq On : 30 Mar 2021 4:42
Operator : JC/MD
Sample : M1832-06
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 49 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IG001

Quant Time: Mar 30 07:10:15 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

