

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032924\
 Data File : VN081624.D
 Acq On : 29 Mar 2024 18:41
 Operator : JC\MD
 Sample : PB159859ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB159859ZHE#02

Quant Time: Mar 30 00:33:44 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	337041	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	628791	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	606737	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	228660	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	235953	54.896	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 109.800%		
35) Dibromofluoromethane	8.171	113	200893	50.624	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 101.240%		
50) Toluene-d8	10.571	98	781015	54.445	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 108.900%		
62) 4-Bromofluorobenzene	12.853	95	249297	49.321	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	= 98.640%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	34604	24.912	ug/l	99
20) Methylene Chloride	5.277	84	13570	2.982	ug/l #	77
43) Isopropyl Acetate	8.688	43	367065	38.724	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032924\
Data File : VN081624.D
Acq On : 29 Mar 2024 18:41
Operator : JC\MD
Sample : PB159859ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB159859ZHE#02

Quant Time: Mar 30 00:33:44 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

