

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032924\
 Data File : VN081625.D
 Acq On : 29 Mar 2024 19:05
 Operator : JC\MD
 Sample : PB159859ZHE#03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB159859ZHE#03

Quant Time: Mar 30 00:34:06 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	332839	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.106	114	637204	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	614287	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	232759	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.583	65	240043	56.553	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	113.100%	
35) Dibromofluoromethane	8.171	113	209426	52.078	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	104.160%	
50) Toluene-d8	10.571	98	790934	54.409	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	108.820%	
62) 4-Bromofluorobenzene	12.853	95	249904	48.788	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	=	97.580%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	36912	26.909	ug/l	94
20) Methylene Chloride	5.283	84	14177	3.155	ug/l #	73
43) Isopropyl Acetate	8.688	43	331410	34.501	ug/l #	84

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032924\
Data File : VN081625.D
Acq On : 29 Mar 2024 19:05
Operator : JC\MD
Sample : PB159859ZHE#03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB159859ZHE#03

Quant Time: Mar 30 00:34:06 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

