

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032924\
 Data File : VN081628.D
 Acq On : 29 Mar 2024 20:16
 Operator : JC\MD
 Sample : PB159859ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB159859ZHE#06

Quant Time: Mar 30 00:35:14 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.229	168	326529	50.000	ug/l	# 0.00
34) 1,4-Difluorobenzene	9.100	114	627587	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.870	117	604806	50.000	ug/l	# 0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	227146	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.582	65	233731	56.130	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	112.260%	
35) Dibromofluoromethane	8.165	113	203273	51.322	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.640%	
50) Toluene-d8	10.570	98	784948	54.824	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	109.640%	
62) 4-Bromofluorobenzene	12.853	95	243399	48.246	ug/l	0.00
Spiked Amount	50.000	Range 64 - 133	Recovery	=	96.500%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	33664	25.015	ug/l	91
20) Methylene Chloride	5.277	84	3443731	781.207	ug/l	85
43) Isopropyl Acetate	8.688	43	317044	33.511	ug/l #	83

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032924\
Data File : VN081628.D
Acq On : 29 Mar 2024 20:16
Operator : JC\MD
Sample : PB159859ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB159859ZHE#06

Quant Time: Mar 30 00:35:14 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

