

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032924\
 Data File : VN081629.D
 Acq On : 29 Mar 2024 20:40
 Operator : JC\MD
 Sample : PB159859ZHE#07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB159859ZHE#07

Quant Time: Mar 30 00:35:37 2024
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
 Quant Title : SW846 8260
 QLast Update : Tue Mar 19 06:00:00 2024
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	357979	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	684505	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	642118	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.794	152	239462	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	250470	54.865	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	= 109.740%	
35) Dibromofluoromethane	8.171	113	214963	49.761	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	= 99.520%	
50) Toluene-d8	10.571	98	841692	53.899	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	= 107.800%	
62) 4-Bromofluorobenzene	12.853	95	259412	47.145	ug/l	0.00
Spiked Amount	50.000	Range	64 - 133	Recovery	= 94.280%	
Target Compounds						Qvalue
16) Acetone	4.424	43	34465	23.360	ug/l	99
20) Methylene Chloride	5.277	84	3779918	782.138	ug/l	85
43) Isopropyl Acetate	8.694	43	324656	31.462	ug/l #	81

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN032924\
Data File : VN081629.D
Acq On : 29 Mar 2024 20:40
Operator : JC\MD
Sample : PB159859ZHE#07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB159859ZHE#07

Quant Time: Mar 30 00:35:37 2024
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031824W.M
Quant Title : SW846 8260
QLast Update : Tue Mar 19 06:00:00 2024
Response via : Initial Calibration

