

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033021\
 Data File : VN066428.D
 Acq On : 30 Mar 2021 18:29
 Operator : JC/MD
 Sample : M1869-01
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT-TB-20210329

Quant Time: Mar 31 02:55:33 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032421W.M
 Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
 QLast Update : Thu Mar 25 05:52:50 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Bromochloromethane	7.110	128	63271	30.00	ug/l	0.00
28) 1,4-Difluorobenzene	8.515	114	322731	30.00	ug/l	0.00
57) Chlorobenzene-d5	11.350	117	271930	30.00	ug/l	0.00

System Monitoring Compounds						
27) 1,2-Dichloroethane-d4	7.949	65	140948	28.98	ug/l	0.00
Spiked Amount	30.000	Range	91 - 110	Recovery	=	96.60%
60) 4-Bromofluorobenzene	12.348	95	100351	23.65	ug/l	0.00
Spiked Amount	30.000	Range	63 - 112	Recovery	=	78.83%
63) Toluene-d8	10.028	98	395273	30.97	ug/l	0.00
Spiked Amount	30.000	Range	91 - 112	Recovery	=	103.23%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033021\
Data File : VN066428.D
Acq On : 30 Mar 2021 18:29
Operator : JC/MD
Sample : M1869-01
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT-TB-20210329

Quant Time: Mar 31 02:55:33 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\624N032421W.M
Quant Title : METHOD 624 VOLATILE ORGANIC ANALYSIS
QLast Update : Thu Mar 25 05:52:50 2021
Response via : Initial Calibration

