

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033021\
 Data File : VN066416.D
 Acq On : 30 Mar 2021 11:23
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 31 02:58:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	84	0.00
2 T	Dichlorodifluoromethane	0.630	0.685	-8.7	90	0.00
3 P	Chloromethane	0.739	0.748	-1.2	94	0.00
4 C	Vinyl Chloride	0.721	0.815	-13.0#	95	0.00
5 T	Bromomethane	0.483	0.533	-10.4	91	-0.01
6 T	Chloroethane	0.449	0.513	-14.3	97	0.00
7 T	Trichlorofluoromethane	0.991	1.052	-6.2	88	0.00
8 T	Diethyl Ether	0.404	0.422	-4.5	90	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.560	0.601	-7.3	90	0.00
10 T	Methyl Iodide	0.794	0.833	-4.9	84	0.00
11 T	Tert butyl alcohol	0.101	0.108	-6.9	83	0.00
12 CM	1,1-Dichloroethene	0.573	0.594	-3.7#	87	0.00
13 T	Acrolein	0.066	0.054	18.2	83	0.00
14 T	Allyl chloride	1.133	1.262	-11.4	90	0.00
15 T	Acrylonitrile	0.283	0.306	-8.1	88	0.00
16 T	Acetone	0.359	0.363	-1.1	86	0.00
17 T	Carbon Disulfide	1.692	1.781	-5.3	89	0.00
18 T	Methyl Acetate	0.938	1.118	-19.2	97	0.00
19 T	Methyl tert-butyl Ether	1.699	1.676	1.4	84	0.00
20 T	Methylene Chloride	0.692	0.722	-4.3	92	0.00
21 T	trans-1,2-Dichloroethene	0.520	0.516	0.8	85	0.00
22 T	Diisopropyl ether	1.768	1.861	-5.3	89	0.00
23 T	Vinyl Acetate	1.505	1.600	-6.3	87	0.00
24 P	1,1-Dichloroethane	0.976	1.005	-3.0	87	0.00
25 T	2-Butanone	0.393	0.421	-7.1	87	0.00
26 T	2,2-Dichloropropane	0.870	0.855	1.7	82	0.00
27 T	cis-1,2-Dichloroethene	0.606	0.614	-1.3	86	0.00
28 T	Bromochloromethane	0.492	0.528	-7.3	90	0.00
29 T	Tetrahydrofuran	0.256	0.284	-10.9	88	0.00
30 C	Chloroform	0.976	1.017	-4.2#	88	0.00
31 T	Cyclohexane	0.920	0.917	0.3	87	0.00
32 T	1,1,1-Trichloroethane	0.857	0.874	-2.0	85	0.00
33 S	1,2-Dichloroethane-d4	0.665	0.664	0.2	83	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	83	0.00
35 S	Dibromofluoromethane	0.331	0.333	-0.6	83	0.00
36 T	1,1-Dichloropropene	0.470	0.484	-3.0	87	0.00
37 T	Ethyl Acetate	0.517	0.561	-8.5	87	0.00
38 T	Carbon Tetrachloride	0.472	0.477	-1.1	84	0.00
39 T	Methylcyclohexane	0.573	0.584	-1.9	84	0.00
40 TM	Benzene	1.419	1.520	-7.1	89	0.00
41 T	Methacrylonitrile	0.257	0.250	2.7	86	0.00
42 TM	1,2-Dichloroethane	0.508	0.529	-4.1	87	0.00
43 T	Isopropyl Acetate	0.884	0.949	-7.4	88	0.00
44 TM	Trichloroethene	0.373	0.383	-2.7	86	0.00
45 C	1,2-Dichloropropane	0.378	0.411	-8.7#	91	0.00
46 T	Dibromomethane	0.248	0.267	-7.7	89	0.00
47 T	Bromodichloromethane	0.512	0.540	-5.5	87	0.00
48 T	Methyl methacrylate	0.420	0.444	-5.7	86	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033021\
 Data File : VN066416.D
 Acq On : 30 Mar 2021 11:23
 Operator : JC/MD
 Sample : VSTDCCC050
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 3 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Mar 31 02:58:27 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.005	0.006	-20.0	95	0.00
50 S	Toluene-d8	1.257	1.299	-3.3	85	0.00
51 T	4-Methyl-2-Pentanone	0.516	0.588	-14.0	90	0.00
52 CM	Toluene	0.871	0.943	-8.3#	89	0.00
53 T	t-1,3-Dichloropropene	0.556	0.591	-6.3	87	0.00
54 T	cis-1,3-Dichloropropene	0.597	0.639	-7.0	87	0.00
55 T	1,1,2-Trichloroethane	0.350	0.386	-10.3	90	0.00
56 T	Ethyl methacrylate	0.537	0.580	-8.0	86	0.00
57 T	1,3-Dichloropropane	0.581	0.641	-10.3	91	0.00
58 T	2-Chloroethyl Vinyl ether	0.048	0.067	-39.6#	110	0.00
59 T	2-Hexanone	0.362	0.422	-16.6	89	0.00
60 T	Dibromochloromethane	0.389	0.419	-7.7	86	0.00
61 T	1,2-Dibromoethane	0.365	0.396	-8.5	88	0.00
62 S	4-Bromofluorobenzene	0.437	0.446	-2.1	83	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	86	0.00
64 T	Tetrachloroethene	0.426	0.404	5.2	82	0.00
65 PM	Chlorobenzene	0.997	1.042	-4.5	88	0.00
66 T	1,1,1,2-Tetrachloroethane	0.384	0.399	-3.9	87	0.00
67 C	Ethyl Benzene	1.809	1.912	-5.7#	89	0.00
68 T	m/p-Xylenes	0.688	0.720	-4.7	88	0.00
69 T	o-Xylene	0.663	0.691	-4.2	88	0.00
70 T	Styrene	1.076	1.141	-6.0	86	0.00
71 P	Bromoform	0.319	0.329	-3.1	83	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	83	0.00
73 T	Isopropylbenzene	3.719	3.810	-2.4	87	0.00
74 T	N-amyl acetate	1.307	1.497	-14.5	85	0.00
75 P	1,1,2,2-Tetrachloroethane	1.184	1.273	-7.5	93	0.00
76 T	1,2,3-Trichloropropane	1.097	1.122	-2.3	87	0.00
77 T	Bromobenzene	0.939	0.969	-3.2	88	0.00
78 T	n-propylbenzene	4.044	4.322	-6.9	87	0.00
79 T	2-Chlorotoluene	2.487	2.531	-1.8	87	0.00
80 T	1,3,5-Trimethylbenzene	3.042	3.129	-2.9	86	0.00
81 T	trans-1,4-Dichloro-2-butene	0.365	0.404	-10.7	87	0.00
82 T	4-Chlorotoluene	2.475	2.588	-4.6	86	0.00
83 T	tert-Butylbenzene	2.619	2.675	-2.1	87	0.00
84 T	1,2,4-Trimethylbenzene	2.956	3.075	-4.0	86	0.00
85 T	sec-Butylbenzene	3.393	3.580	-5.5	86	0.00
86 T	p-Isopropyltoluene	3.039	3.197	-5.2	85	0.00
87 T	1,3-Dichlorobenzene	1.576	1.659	-5.3	87	0.00
88 T	1,4-Dichlorobenzene	1.636	1.691	-3.4	86	0.00
89 T	n-Butylbenzene	2.510	2.666	-6.2	85	0.00
90 T	Hexachloroethane	0.548	0.569	-3.8	87	0.00
91 T	1,2-Dichlorobenzene	1.549	1.640	-5.9	87	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.227	0.240	-5.7	85	0.00
93 T	1,2,4-Trichlorobenzene	0.961	1.009	-5.0	86	0.00
94 T	Hexachlorobutadiene	0.524	0.542	-3.4	87	0.00
95 T	Naphthalene	2.899	2.912	-0.4	87	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033021\
Data File : VN066416.D
Acq On : 30 Mar 2021 11:23
Operator : JC/MD
Sample : VSTDCCC050
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 3 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Mar 31 02:58:27 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.966	0.994	-2.9	88	0.00

(#) = Out of Range SPCC's out = 0 CCC's out = 6