

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
 Data File : VN071655.D
 Acq On : 30 Mar 2022 13:35
 Operator : JC\MD
 Sample : VN0330MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0330MBL01

Quant Time: Mar 31 01:32:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 19:18:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	594516	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	978852	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	956050	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	357462	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	334651	49.307	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.620%
35) Dibromofluoromethane	8.016	113	281578	48.885	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.780%
50) Toluene-d8	10.439	98	1202158	51.019	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.040%
62) 4-Bromofluorobenzene	12.727	95	397232	52.126	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.260%
Target Compounds						
					Qvalue	
3) Chloromethane	2.304	50	1647	0.275	ug/l	95
4) Vinyl Chloride	2.446	62	133	0.020	ug/l #	43
5) Bromomethane	2.881	94	2079	0.486	ug/l #	80
6) Chloroethane	3.122	64	2931	0.619	ug/l #	58
10) Methyl Iodide	4.440	142	1269	0.176	ug/l #	71
13) Acrolein	4.040	56	661	0.492	ug/l	94
15) Acrylonitrile	5.557	53	953	0.298	ug/l	88
16) Acetone	4.298	43	3168	1.092	ug/l #	86
17) Carbon Disulfide	4.522	76	1096	0.097	ug/l #	75
18) Methyl Acetate	4.857	43	4060	Below Cal	#	48
20) Methylene Chloride	5.087	84	531	0.087	ug/l #	68
21) trans-1,2-Dichloroethene	5.587	96	155	0.029	ug/l #	1
23) Vinyl Acetate	6.439	43	516	0.037	ug/l #	68
25) 2-Butanone	7.339	43	1563	0.374	ug/l #	45
27) cis-1,2-Dichloroethene	7.328	96	57	0.009	ug/l #	1
29) Tetrahydrofuran	7.704	42	714	0.264	ug/l #	80
30) Chloroform	7.804	83	56	0.005	ug/l #	18
31) Cyclohexane	8.086	56	10396	1.126	ug/l #	14
36) 1,1-Dichloropropene	8.081	75	40405	5.023	ug/l #	48
37) Ethyl Acetate	7.422	43	161	0.018	ug/l #	67
38) Carbon Tetrachloride	8.081	117	52444	6.653	ug/l #	18
40) Benzene	8.451	78	455	0.018	ug/l #	51
41) Methacrylonitrile	7.651	41	58	0.013	ug/l #	16
42) 1,2-Dichloroethane	8.522	62	945	0.108	ug/l #	74
43) Isopropyl Acetate	8.639	43	12614	0.861	ug/l #	53
44) Trichloroethene	9.210	130	160	0.024	ug/l	80
47) Bromodichloromethane	9.757	83	65	0.008	ug/l #	26
48) Methyl methacrylate	9.557	41	163	0.027	ug/l #	57
51) 4-Methyl-2-Pentanone	10.322	43	2979	0.354	ug/l #	87
52) Toluene	10.504	92	629	0.039	ug/l #	75
53) t-1,3-Dichloropropene	10.722	75	419	0.047	ug/l #	1
54) cis-1,3-Dichloropropene	10.175	75	115	0.012	ug/l #	36
55) 1,1,2-Trichloroethane	10.892	97	125	0.019	ug/l #	16
56) Ethyl methacrylate	10.751	69	532	0.056	ug/l #	52
57) 1,3-Dichloropropane	11.045	76	213	0.019	ug/l #	42
58) 2-Chloroethyl Vinyl ether	10.033	63	66	16.988	ug/l #	44

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
 Data File : VN071655.D
 Acq On : 30 Mar 2022 13:35
 Operator : JC\MD
 Sample : VN0330MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0330MBL01

Quant Time: Mar 31 01:32:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 19:18:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)


59) 2-Hexanone	11.086	43	7083	1.195	ug/l	95
60) Dibromochloromethane	11.227	129	115	0.018	ug/l #	11
61) 1,2-Dibromoethane	11.339	107	56	0.009	ug/l #	55
64) Tetrachloroethene	10.980	164	309	0.049	ug/l #	50
65) Chlorobenzene	11.769	112	1914	0.100	ug/l	96
67) Ethyl Benzene	11.839	91	1906	0.059	ug/l #	82
68) m/p-Xylenes	11.945	106	1514	0.122	ug/l	75
69) o-Xylene	12.274	106	517	0.042	ug/l	42
70) Styrene	12.298	104	1212	0.065	ug/l	100
73) Isopropylbenzene	12.580	105	1897	0.063	ug/l	75
74) N-amyl acetate	12.392	43	4064	0.471	ug/l	92
75) 1,1,2,2-Tetrachloroethane	12.821	83	919	0.091	ug/l #	68
76) 1,2,3-Trichloropropane	12.727	75	196860	23.773	ug/l #	1
77) Bromobenzene	12.868	156	1010	0.134	ug/l	90
78) n-propylbenzene	12.916	91	3823	0.121	ug/l	93
79) 2-Chlorotoluene	13.004	91	2315	0.114	ug/l	91
80) 1,3,5-Trimethylbenzene	13.051	105	1748	0.073	ug/l	96
81) trans-1,4-Dichloro-2-b...	12.727	75	196860	73.671	ug/l #	11
82) 4-Chlorotoluene	13.104	91	3290	0.175	ug/l	98
83) tert-Butylbenzene	13.327	119	1741	0.080	ug/l	90
84) 1,2,4-Trimethylbenzene	13.357	105	2467	0.106	ug/l	98
85) sec-Butylbenzene	13.498	105	3059	0.108	ug/l	95
86) p-Isopropyltoluene	13.610	119	2997	0.133	ug/l	98
87) 1,3-Dichlorobenzene	13.610	146	2795	0.232	ug/l	95
88) 1,4-Dichlorobenzene	13.610	146	2795	0.238	ug/l	97
89) n-Butylbenzene	13.939	91	5980	0.358	ug/l	99
90) Hexachloroethane	13.945	117	133	0.032	ug/l #	9
91) 1,2-Dichlorobenzene	13.986	146	2756	0.233	ug/l	95
92) 1,2-Dibromo-3-Chloropr...	14.604	75	318	0.204	ug/l	57
93) 1,2,4-Trichlorobenzene	15.257	180	3302	2.642	ug/l	82
94) Hexachlorobutadiene	15.357	225	679	0.195	ug/l	86
95) Naphthalene	15.504	128	11479	4.454	ug/l #	95
96) 1,2,3-Trichlorobenzene	15.692	180	3709	3.055	ug/l	81


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
 Data File : VN071655.D
 Acq On : 30 Mar 2022 13:35
 Operator : JC\MD
 Sample : VN0330MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0330MBL01

Quant Time: Mar 31 01:32:18 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 19:18:20 2022
 Response via : Initial Calibration









#1
Pentafluorobenzene
Concen: 50.000 ug/l
RT: 8.081 min Scan# 1
Delta R.T. -0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion:168 Resp: 594516
Ion Ratio Lower Upper
168 100
99 52.6 42.3 63.5



#3
Chloromethane
Concen: 0.275 ug/l
RT: 2.304 min Scan# 54
Delta R.T. 0.005 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 50 Resp: 1647
Ion Ratio Lower Upper
50 100
52 29.4 25.7 38.5









#4
 Vinyl Chloride
 Concen: 0.020 ug/l
 RT: 2.446 min Scan# 71
 Delta R.T. -0.000 min
 Lab File: VN071655.D
 Acq: 30 Mar 2022 13:35




Instrument :
 MSVOA_N
 ClientSampleId :
 VN0330MBL01


Tgt Ion: 62 Resp: 133
 Ion Ratio Lower Upper
 62 100
 64 0.0 25.0 37.4#



#5
 Bromomethane
 Concen: 0.486 ug/l
 RT: 2.881 min Scan# 152
 Delta R.T. 0.035 min
 Lab File: VN071655.D
 Acq: 30 Mar 2022 13:35

Tgt Ion: 94 Resp: 2079
 Ion Ratio Lower Upper
 94 100
 96 75.7 76.0 114.0#








#6
Chloroethane
Concen: 0.619 ug/l
RT: 3.122 min Scan# 193
Delta R.T. 0.112 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 64 Resp: 2931
Ion Ratio Lower Upper
64 100
66 9.3 26.2 39.4#



#10
Methyl Iodide
Concen: 0.176 ug/l
RT: 4.440 min Scan# 417
Delta R.T. 0.018 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 142 Resp: 1269
Ion Ratio Lower Upper
142 100
127 22.3 31.3 46.9#
141 0.0 11.2 16.8#









#13
Acrolein
Concen: 0.492 ug/l
RT: 4.040 min Scan# 349
Delta R.T. -0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 56 Resp: 661
Ion Ratio Lower Upper
56 100
55 63.5 54.8 82.2



#15
Acrylonitrile
Concen: 0.298 ug/l
RT: 5.557 min Scan# 607
Delta R.T. 0.011 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 53 Resp: 953
Ion Ratio Lower Upper
53 100
52 70.0 66.4 99.6
51 31.5 28.5 42.7









#16
Acetone
Concen: 1.092 ug/l
RT: 4.298 min Scan# 393
Delta R.T. 0.023 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 43 Resp: 3168
Ion Ratio Lower Upper
43 100
58 24.6 26.0 39.0#



#17
Carbon Disulfide
Concen: 0.097 ug/l
RT: 4.522 min Scan# 431
Delta R.T. -0.012 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 76 Resp: 1096
Ion Ratio Lower Upper
76 100
78 0.0 7.2 10.8#





#18

Methyl Acetate

Concen: Below Cal

RT: 4.857 min Scan# 41

Delta R.T. 0.011 min

Lab File: VN071655.D

Acq: 30 Mar 2022 13:35

Instrument :

MSVOA_N

ClientSampleId :





VN0330MBL01

Tgt Ion: 43 Resp: 4060

Ion Ratio Lower Upper

43 100

74 0.0 21.6 32.4#



#20

Methylene Chloride

Concen: 0.087 ug/l

RT: 5.087 min Scan# 527

Delta R.T. -0.006 min

Lab File: VN071655.D

Acq: 30 Mar 2022 13:35

Tgt Ion: 84 Resp: 531




Ion Ratio Lower Upper


84 100

49 103.0 88.3 132.5

51 0.0 28.2 42.2#

86 109.6 50.6 76.0#









#21
trans-1,2-Dichloroethene
Concen: 0.029 ug/l
RT: 5.587 min Scan# 61
Delta R.T. -0.012 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 96 Resp: 155
Ion Ratio Lower Upper
96 100
61 0.0 105.8 158.8#
98 0.0 50.3 75.5#



#23
Vinyl Acetate
Concen: 0.037 ug/l
RT: 6.439 min Scan# 757
Delta R.T. 0.011 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 43 Resp: 516
Ion Ratio Lower Upper
43 100
86 0.0 10.0 15.0#





#25

2-Butanone

Concen: 0.374 ug/l

RT: 7.339 min Scan# 91

Delta R.T. 0.011 min

Lab File: VN071655.D

Acq: 30 Mar 2022 13:35

Instrument :

MSVOA_N

ClientSampleId :





VN0330MBL01

Tgt Ion: 43 Resp: 1563

Ion Ratio Lower Upper

43 100

72 0.0 23.7 35.5#



#27

cis-1,2-Dichloroethene

Concen: 0.009 ug/l

RT: 7.328 min Scan# 908

Delta R.T. 0.006 min

Lab File: VN071655.D

Acq: 30 Mar 2022 13:35




Tgt Ion: 96 Resp: 57


Ion Ratio Lower Upper

96 100

61 640.4 0.0 280.6#

98 0.0 0.0 128.4





#29

Tetrahydrofuran

Concen: 0.264 ug/l

RT: 7.704 min Scan# 91

Delta R.T. 0.029 min

Lab File: VN071655.D

Acq: 30 Mar 2022 13:35

Instrument :

MSVOA_N

ClientSampleId :

VN0330MBL01





Tgt Ion: 42 Resp: 714

Ion Ratio Lower Upper

42 100

72 27.7 39.7 59.5#

71 41.3 36.6 55.0



#30

Chloroform

Concen: 0.005 ug/l

RT: 7.804 min Scan# 989

Delta R.T. -0.006 min

Lab File: VN071655.D




Acq: 30 Mar 2022 13:35


Tgt Ion: 83 Resp: 56

Ion Ratio Lower Upper

83 100

85 0.0 51.8 77.8#





#31

Cyclohexane

Concen: 1.126 ug/l

RT: 8.086 min Scan# 1037

Delta R.T. -0.006 min

Lab File: VN071655.D

Acq: 30 Mar 2022 13:35

Instrument :

MSVOA_N

ClientSampleId :

VN0330MBL01





Tgt Ion: 56 Resp: 10396

Ion Ratio Lower Upper

56 100

69 169.1 27.8 41.8#

84 128.2 76.9 115.3#



#33

1,2-Dichloroethane-d4

Concen: 49.307 ug/l

RT: 8.433 min Scan# 1096

Delta R.T. -0.001 min

Lab File: VN071655.D




Acq: 30 Mar 2022 13:35


Tgt Ion: 65 Resp: 334651

Ion Ratio Lower Upper

65 100

67 51.7 0.0 104.8









#34
1,4-Difluorobenzene
Concen: 50.000 ug/l
RT: 8.963 min Scan# 1186
Delta R.T. -0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion:114 Resp: 978852
Ion Ratio Lower Upper
114 100
63 18.6 0.0 40.0
88 15.8 0.0 32.6



#35
Dibromofluoromethane
Concen: 48.885 ug/l
RT: 8.016 min Scan# 1025
Delta R.T. -0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion:113 Resp: 281578
Ion Ratio Lower Upper
113 100
111 102.0 81.4 122.0
192 22.3 17.0 25.6





#36

1,1-Dichloropropene

Concen: 5.023 ug/l

RT: 8.081 min Scan# 1059

Delta R.T. -0.135 min

Lab File: VN071655.D

Acq: 30 Mar 2022 13:35

Instrument :

MSVOA_N

ClientSampleId :

VN0330MBL01





Tgt Ion: 75 Resp: 40405

Ion Ratio Lower Upper

75 100

110 8.4 18.3 54.8#

77 0.0 25.1 37.7#



#37

Ethyl Acetate

Concen: 0.018 ug/l

RT: 7.422 min Scan# 924

Delta R.T. 0.018 min

Lab File: VN071655.D

Acq: 30 Mar 2022 13:35




Tgt Ion: 43 Resp: 161


Ion Ratio Lower Upper

43 100

61 0.0 11.8 17.6#

70 0.0 9.3 13.9#









#38
Carbon Tetrachloride
Concen: 6.653 ug/l
RT: 8.081 min Scan# 1099
Delta R.T. -0.123 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion:117 Resp: 52444
Ion Ratio Lower Upper
117 100
119 7.6 75.8 113.8#
121 0.0 24.6 37.0#



#40
Benzene
Concen: 0.018 ug/l
RT: 8.451 min Scan# 1099
Delta R.T. -0.006 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 78 Resp: 455
Ion Ratio Lower Upper
78 100
77 0.0 19.4 29.0#









#41
Methacrylonitrile
Concen: 0.013 ug/l
RT: 7.651 min Scan# 90
Delta R.T. 0.023 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion:	41	Resp:	58
Ion	Ratio	Lower	Upper
41	100		
39	0.0	48.6	73.0#
67	0.0	73.4	110.2#
52	0.0	27.7	41.5#



#42
1,2-Dichloroethane
Concen: 0.108 ug/l
RT: 8.522 min Scan# 1111
Delta R.T. -0.006 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion:	62	Resp:	945
Ion	Ratio	Lower	Upper
62	100		
98	0.0	0.0	19.4





#43

Isopropyl Acetate

Concen: 0.861 ug/l

RT: 8.639 min Scan# 1116

Delta R.T. 0.088 min

Lab File: VN071655.D

Acq: 30 Mar 2022 13:35

Instrument :

MSVOA_N

ClientSampleId :

VN0330MBL01


Tgt Ion: 43 Resp: 12614

Ion Ratio Lower Upper

43 100

61 0.0 21.9 32.9#

87 0.0 12.0 18.0#



Abundance

3000



8.639

2000

1000

0

Time-->



#44

Trichloroethene

Concen: 0.024 ug/l

RT: 9.210 min Scan# 1228

Delta R.T. -0.000 min

Lab File: VN071655.D


Acq: 30 Mar 2022 13:35

Tgt Ion: 130 Resp: 160

Ion Ratio Lower Upper

130 100

95 73.8 0.0 186.8



Abundance

300


9.210


200

100

0

Time-->









#47
Bromodichloromethane
Concen: 0.008 ug/l
RT: 9.757 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 83 Resp: 65
Ion Ratio Lower Upper
83 100
85 0.0 50.5 75.7#
127 0.0 6.8 10.2#



#48
Methyl methacrylate
Concen: 0.027 ug/l
RT: 9.557 min Scan# 1287
Delta R.T. 0.006 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 41 Resp: 163
Ion Ratio Lower Upper
41 100
69 40.5 74.7 112.1#
39 40.5 45.6 68.4#









#50
Toluene-d8
Concen: 51.019 ug/l
RT: 10.439 min Scan# 1437
Delta R.T. 0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 98 Resp: 1202158
Ion Ratio Lower Upper
98 100
100 64.4 52.6 78.8



#51
4-Methyl-2-Pentanone
Concen: 0.354 ug/l
RT: 10.322 min Scan# 1417
Delta R.T. -0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 43 Resp: 2979
Ion Ratio Lower Upper
43 100
58 34.0 34.1 51.1#









#52
Toluene
Concen: 0.039 ug/l
RT: 10.504 min Scan# 1448
Delta R.T. 0.006 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 92 Resp: 629
Ion Ratio Lower Upper
92 100
91 208.4 138.6 207.8#



#53
t-1,3-Dichloropropene
Concen: 0.047 ug/l
RT: 10.722 min Scan# 1485
Delta R.T. 0.012 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 75 Resp: 419
Ion Ratio Lower Upper
75 100
77 111.2 25.6 38.4#









#54
cis-1,3-Dichloropropene
Concen: 0.012 ug/l
RT: 10.175 min Scan# 1115
Delta R.T. -0.011 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 75 Resp: 115
Ion Ratio Lower Upper
75 100
77 0.0 25.2 37.8#
39 0.0 35.7 53.5#



#55
1,1,2-Trichloroethane
Concen: 0.019 ug/l
RT: 10.892 min Scan# 1514
Delta R.T. 0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 97 Resp: 125
Ion Ratio Lower Upper
97 100
83 0.0 66.2 99.4#
85 0.0 43.9 65.9#
99 0.0 50.6 76.0#









#56
Ethyl methacrylate
Concen: 0.056 ug/l
RT: 10.751 min Scan# 1490
Delta R.T. -0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 69 Resp: 532
Ion Ratio Lower Upper
69 100
41 91.2 47.8 71.6#
39 0.0 27.3 40.9#



#57
1,3-Dichloropropane
Concen: 0.019 ug/l
RT: 11.045 min Scan# 1540
Delta R.T. 0.006 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 76 Resp: 213
Ion Ratio Lower Upper
76 100
78 0.0 26.2 39.2#









#58
2-Chloroethyl Vinyl ether
Concen: 16.988 ug/l
RT: 10.033 min Scan# 11
Delta R.T. -0.001 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 63 Resp: 66
Ion Ratio Lower Upper
63 100
106 0.0 24.6 36.8#



#59
2-Hexanone
Concen: 1.195 ug/l
RT: 11.086 min Scan# 1547
Delta R.T. 0.005 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 43 Resp: 7083
Ion Ratio Lower Upper
43 100
58 55.4 29.7 89.1









#60
Dibromochloromethane
Concen: 0.018 ug/l
RT: 11.227 min Scan# 11
Delta R.T. -0.006 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion:129 Resp: 115
Ion Ratio Lower Upper
129 100
127 0.0 38.8 116.4#



#61
1,2-Dibromoethane
Concen: 0.009 ug/l
RT: 11.339 min Scan# 1590
Delta R.T. 0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion:107 Resp: 56
Ion Ratio Lower Upper
107 100
109 137.5 75.0 112.6#









#62
4-Bromofluorobenzene
Concen: 52.126 ug/l
RT: 12.727 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 95 Resp: 397232
Ion Ratio Lower Upper
95 100
174 85.4 0.0 167.6
176 83.5 0.0 163.4



#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.739 min Scan# 1658
Delta R.T. 0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 117 Resp: 956050
Ion Ratio Lower Upper
117 100
82 55.6 44.2 66.2
119 31.9 25.4 38.0









#64
Tetrachloroethene
Concen: 0.049 ug/l
RT: 10.980 min Scan# 11
Delta R.T. 0.005 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion:164 Resp: 309
Ion Ratio Lower Upper
164 100
166 0.0 104.2 156.2#
129 97.5 73.9 110.9
131 90.5 69.4 104.2



#65
Chlorobenzene
Concen: 0.100 ug/l
RT: 11.769 min Scan# 1663
Delta R.T. 0.006 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion:112 Resp: 1914
Ion Ratio Lower Upper
112 100
114 34.3 25.4 38.2









#67
Ethyl Benzene
Concen: 0.059 ug/l
RT: 11.839 min Scan# 1675
Delta R.T. 0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 91 Resp: 1906
Ion Ratio Lower Upper
91 100
106 20.9 24.9 37.3#



#68
m/p-Xylenes
Concen: 0.122 ug/l
RT: 11.945 min Scan# 1693
Delta R.T. -0.006 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 106 Resp: 1514
Ion Ratio Lower Upper
106 100
91 235.9 158.9 238.3









#69
o-Xylene
Concen: 0.042 ug/l
RT: 12.274 min Scan# 1749
Delta R.T. -0.001 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion:106 Resp: 517
Ion Ratio Lower Upper
106 100
91 302.7 106.0 318.0



#70
Styrene
Concen: 0.065 ug/l
RT: 12.298 min Scan# 1753
Delta R.T. 0.012 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion:104 Resp: 1212
Ion Ratio Lower Upper
104 100
78 49.3 39.5 59.3
103 54.7 44.0 66.0









#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.668 min Scan# 1986
Delta R.T. -0.001 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion:152 Resp: 357462
Ion Ratio Lower Upper
152 100
115 55.8 28.9 86.7
150 155.1 0.0 356.2



#73
Isopropylbenzene
Concen: 0.063 ug/l
RT: 12.580 min Scan# 1801
Delta R.T. 0.005 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion:105 Resp: 1897
Ion Ratio Lower Upper
105 100
120 13.8 13.5 40.4









#74
N-amy1 acetate
Concen: 0.471 ug/l
RT: 12.392 min Scan# 1769
Delta R.T. 0.012 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 43	Resp: 4064
Ion Ratio	Lower Upper
43 100	
70 42.4	39.0 58.4
55 24.5	21.1 31.7
61 21.2	21.1 31.7



#75
1,1,2,2-Tetrachloroethane
Concen: 0.091 ug/l
RT: 12.821 min Scan# 1842
Delta R.T. -0.001 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 83	Resp:	919	
Ion Ratio	Lower	Upper	
83 100			
131 0.0	5.3	15.9#	
85 39.3	32.3	96.9	









#76
1,2,3-Trichloropropane
Concen: 23.773 ug/l
RT: 12.727 min Scan# 11
Delta R.T. -0.147 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 75 Resp: 196860
Ion Ratio Lower Upper
75 100
77 1.2 98.5 295.5#



#77
Bromobenzene
Concen: 0.134 ug/l
RT: 12.868 min Scan# 1850
Delta R.T. 0.017 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 156 Resp: 1010
Ion Ratio Lower Upper
156 100
77 222.1 100.5 301.5
158 101.2 48.9 146.6









#78
n-propylbenzene
Concen: 0.121 ug/l
RT: 12.916 min Scan# 1858
Delta R.T. -0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 91 Resp: 3823
Ion Ratio Lower Upper
91 100
120 27.3 11.9 35.5



#79
2-Chlorotoluene
Concen: 0.114 ug/l
RT: 13.004 min Scan# 1873
Delta R.T. -0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 91 Resp: 2315
Ion Ratio Lower Upper
91 100
126 29.6 17.3 51.9









#80
1,3,5-Trimethylbenzene
Concen: 0.073 ug/l
RT: 13.051 min Scan# 11
Delta R.T. -0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion:105 Resp: 1748
Ion Ratio Lower Upper
105 100
120 52.6 24.9 74.7



#81
trans-1,4-Dichloro-2-butene
Concen: 73.671 ug/l
RT: 12.727 min Scan# 1826
Delta R.T. 0.105 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 75 Resp: 196860
Ion Ratio Lower Upper
75 100
53 0.1 78.0 117.0#
89 0.0 37.2 55.8#









#82
4-Chlorotoluene
Concen: 0.175 ug/l
RT: 13.104 min Scan# 1889
Delta R.T. 0.006 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 91 Resp: 3290
Ion Ratio Lower Upper
91 100
126 35.4 17.2 51.6



#83
tert-Butylbenzene
Concen: 0.080 ug/l
RT: 13.327 min Scan# 1928
Delta R.T. 0.011 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 119 Resp: 1741
Ion Ratio Lower Upper
119 100
91 71.4 31.6 94.8
134 32.8 13.7 41.1









#84
1,2,4-Trimethylbenzene
Concen: 0.106 ug/l
RT: 13.357 min Scan# 1933
Delta R.T. -0.006 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion:105 Resp: 2467
Ion Ratio Lower Upper
105 100
120 48.0 23.3 69.8



#85
sec-Butylbenzene
Concen: 0.108 ug/l
RT: 13.498 min Scan# 1957
Delta R.T. -0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion:105 Resp: 3059
Ion Ratio Lower Upper
105 100
134 18.4 10.4 31.1









#86
p-Isopropyltoluene
Concen: 0.133 ug/l
RT: 13.610 min Scan# 1976
Delta R.T. -0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01

Tgt Ion:119 Resp: 2997




Ion	Ratio	Lower	Upper
119	100		
134	25.4	13.3	39.8
91	23.3	11.3	33.9




#87
1,3-Dichlorobenzene
Concen: 0.232 ug/l
RT: 13.610 min Scan# 1976
Delta R.T. -0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion:146 Resp: 2795

Ion	Ratio	Lower	Upper
146	100		
111	33.3	20.0	60.0
148	63.3	32.0	96.0









#88
1,4-Dichlorobenzene
Concen: 0.238 ug/l
RT: 13.610 min Scan# 1989
Delta R.T. -0.076 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01

Tgt Ion:146 Resp: 2795




Ion	Ratio	Lower	Upper
146	100		
111	33.3	19.1	57.1
148	63.3	31.4	94.3




#89
n-Butylbenzene
Concen: 0.358 ug/l
RT: 13.939 min Scan# 2032
Delta R.T. -0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 91 Resp: 5980

Ion	Ratio	Lower	Upper
91	100		
92	53.7	26.9	80.7
134	25.1	13.6	40.8









#90
Hexachloroethane
Concen: 0.032 ug/l
RT: 13.945 min Scan# 2077
Delta R.T. -0.259 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion:117 Resp: 133
Ion Ratio Lower Upper
117 100
201 0.0 40.6 122.0#



#91
1,2-Dichlorobenzene
Concen: 0.233 ug/l
RT: 13.986 min Scan# 2040
Delta R.T. 0.006 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion:146 Resp: 2756
Ion Ratio Lower Upper
146 100
111 44.6 20.9 62.8
148 68.4 32.1 96.3









#92
1,2-Dibromo-3-Chloropropane
Concen: 0.204 ug/l
RT: 14.604 min Scan# 2145
Delta R.T. 0.006 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion: 75 Resp: 318
Ion Ratio Lower Upper
75 100
155 134.0 44.9 134.5
157 161.0 59.0 176.8



#93
1,2,4-Trichlorobenzene
Concen: 2.642 ug/l
RT: 15.257 min Scan# 2256
Delta R.T. -0.000 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35

Tgt Ion: 180 Resp: 3302
Ion Ratio Lower Upper
180 100
182 115.8 47.5 142.5
145 32.1 14.5 43.5









#94
Hexachlorobutadiene
Concen: 0.195 ug/l
RT: 15.357 min Scan# 21
Delta R.T. -0.006 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35




Instrument :
MSVOA_N
ClientSampleId :
VN0330MBL01


Tgt Ion:	225	Resp:	679
Ion	Ratio	Lower	Upper
225	100		
223	58.2	32.4	97.2
227	48.5	31.8	95.4



#95
Naphthalene
Concen: 4.454 ug/l
RT: 15.504 min Scan# 2298
Delta R.T. 0.006 min
Lab File: VN071655.D
Acq: 30 Mar 2022 13:35


Tgt Ion:	128	Resp:	11479
Ion	Ratio	Lower	Upper
128	100		
127	16.6	10.3	15.5#
129	10.9	8.6	13.0





#96
 1,2,3-Trichlorobenzene
 Concen: 3.055 ug/l
 RT: 15.692 min Scan# 21
 Delta R.T. -0.000 min
 Lab File: VN071655.D
 Acq: 30 Mar 2022 13:35

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0330MBL01



Tgt Ion:	180	Resp:	3709
Ion	Ratio	Lower	Upper
180	100		
182	114.4	47.8	143.3
145	40.2	15.5	46.5

