

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
 Data File : VN071653.D
 Acq On : 30 Mar 2022 12:27
 Operator : JC\MD
 Sample : IBLK
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 5 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 IBLK

Quant Time: Mar 31 01:31:16 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 19:18:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	533269	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	873355	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	838402	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	305278	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	299312	49.165	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.340%
35) Dibromofluoromethane	8.016	113	247002	48.062	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.120%
50) Toluene-d8	10.439	98	1077453	51.251	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.500%
62) 4-Bromofluorobenzene	12.727	95	348594	51.270	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	102.540%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
Data File : VN071653.D
Acq On : 30 Mar 2022 12:27
Operator : JC\MD
Sample : IBLK
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 5 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
IBLK

Quant Time: Mar 31 01:31:16 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 19:18:20 2022
Response via : Initial Calibration

