

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
 Data File : VN071656.D
 Acq On : 30 Mar 2022 13:59
 Operator : JC\MD
 Sample : VN0330WBL01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0330WBL01

Quant Time: Mar 31 01:32:37 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 19:18:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	625211	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1051096	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1042295	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	387047	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	355909	49.865	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.720%
35) Dibromofluoromethane	8.016	113	299987	48.502	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.000%
50) Toluene-d8	10.439	98	1280760	50.619	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.240%
62) 4-Bromofluorobenzene	12.727	95	436278	53.315	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	106.640%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
Data File : VN071656.D
Acq On : 30 Mar 2022 13:59
Operator : JC\MD
Sample : VN0330WBL01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0330WBL01

Quant Time: Mar 31 01:32:37 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 19:18:20 2022
Response via : Initial Calibration

