

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
 Data File : VN071669.D
 Acq On : 30 Mar 2022 19:20
 Operator : JC\MD
 Sample : N2075-01
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE104D2-20220321

Quant Time: Mar 31 01:36:46 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 19:18:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	625178	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1036824	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1007003	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	327187	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	351506	49.250	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.500%
35) Dibromofluoromethane	8.016	113	293615	48.125	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.240%
50) Toluene-d8	10.439	98	1286857	51.560	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.120%
62) 4-Bromofluorobenzene	12.727	95	393385	48.735	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	97.480%
Target Compounds						
					Qvalue	
21) trans-1,2-Dichloroethene	5.593	96	3048	0.538	ug/l #	70
27) cis-1,2-Dichloroethene	7.328	96	128187	18.453	ug/l	86
30) Chloroform	7.816	83	10071	0.891	ug/l	97
42) 1,2-Dichloroethane	8.528	62	7045	0.757	ug/l	97
44) Trichloroethene	9.216	130	1035240	149.301	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
Data File : VN071669.D
Acq On : 30 Mar 2022 19:20
Operator : JC\MD
Sample : N2075-01
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE104D2-20220321

Quant Time: Mar 31 01:36:46 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 19:18:20 2022
Response via : Initial Calibration

