

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
 Data File : VN071670.D
 Acq On : 30 Mar 2022 19:44
 Operator : JC\MD
 Sample : N2075-11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 25 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW179D1-20220322

Quant Time: Mar 31 01:37:05 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 19:18:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	661357	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1110980	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1103886	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	392686	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	375538	49.739	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.480%
35) Dibromofluoromethane	8.016	113	316579	48.425	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.860%
50) Toluene-d8	10.439	98	1372101	51.306	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.620%
62) 4-Bromofluorobenzene	12.727	95	455213	52.631	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.260%
Target Compounds						
					Qvalue	
27) cis-1,2-Dichloroethene	7.322	96	70225	9.556	ug/l	90
44) Trichloroethene	9.210	130	699829	94.192	ug/l	98
64) Tetrachloroethene	10.974	164	69038	9.471	ug/l	98

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
Data File : VN071670.D
Acq On : 30 Mar 2022 19:44
Operator : JC\MD
Sample : N2075-11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 25 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW179D1-20220322

Quant Time: Mar 31 01:37:05 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 19:18:20 2022
Response via : Initial Calibration

