

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
 Data File : VN071679.D
 Acq On : 31 Mar 2022 00:08
 Operator : JC\MD
 Sample : N2075-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 36 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TT161S1-20220321

Quant Time: Mar 31 01:40:09 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 19:18:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	674821	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1128517	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1136300	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	418603	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	384034	49.849	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.700%
35) Dibromofluoromethane	8.016	113	327861	49.372	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.740%
50) Toluene-d8	10.439	98	1402717	51.636	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.280%
62) 4-Bromofluorobenzene	12.727	95	476439	54.229	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	108.460%
Target Compounds						
					Qvalue	
16) Acetone	4.304	43	8116	2.465	ug/l	94
19) Methyl tert-butyl Ether	5.610	73	10600	0.512	ug/l	94
24) 1,1-Dichloroethane	6.387	63	26987	2.336	ug/l	98
30) Chloroform	7.822	83	3240	0.266	ug/l #	68
44) Trichloroethene	9.210	130	8916	1.181	ug/l	82

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
Data File : VN071679.D
Acq On : 31 Mar 2022 00:08
Operator : JC\MD
Sample : N2075-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 36 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TT161S1-20220321

Quant Time: Mar 31 01:40:09 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 19:18:20 2022
Response via : Initial Calibration

