

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
 Data File : VN071680.D
 Acq On : 31 Mar 2022 00:32
 Operator : JC\MD
 Sample : N2075-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 GM38-RW3-MW1-20220321

Quant Time: Mar 31 01:40:32 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 19:18:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	633963	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1059725	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1046919	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	348376	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	357318	49.371	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	98.740%
35) Dibromofluoromethane	8.022	113	301710	48.383	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.760%
50) Toluene-d8	10.439	98	1307004	51.236	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.480%
62) 4-Bromofluorobenzene	12.727	95	412358	49.982	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.960%
Target Compounds						
					Qvalue	
16) Acetone	4.293	43	5724	1.851	ug/l #	84
44) Trichloroethene	9.210	130	115029	16.231	ug/l	95
64) Tetrachloroethene	10.969	164	8447	1.222	ug/l	92

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
Data File : VN071680.D
Acq On : 31 Mar 2022 00:32
Operator : JC\MD
Sample : N2075-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
GM38-RW3-MW1-20220321

Quant Time: Mar 31 01:40:32 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 19:18:20 2022
Response via : Initial Calibration

