

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
 Data File : VN071684.D
 Acq On : 31 Mar 2022 02:08
 Operator : JC\MD
 Sample : N2075-07
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 41 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE109D3-20220322

Quant Time: Mar 31 07:46:10 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	625996	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1068138	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1033565	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	345394	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	361265	50.551	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.100%
35) Dibromofluoromethane	8.022	113	308148	49.026	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.060%
50) Toluene-d8	10.439	98	1324977	51.531	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.060%
62) 4-Bromofluorobenzene	12.727	95	413952	49.780	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	99.560%
Target Compounds						
						Qvalue
9) 1,1,2-Trichlorotrifluo...	4.216	101	5291	0.974	ug/l	98
16) Acetone	4.293	43	5735	1.878	ug/l #	85
38) Carbon Tetrachloride	8.198	117	2673	0.311	ug/l	92
44) Trichloroethene	9.216	130	230977	32.335	ug/l	99
64) Tetrachloroethene	10.975	164	4238	0.621	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
Data File : VN071684.D
Acq On : 31 Mar 2022 02:08
Operator : JC\MD
Sample : N2075-07
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 41 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE109D3-20220322

Quant Time: Mar 31 07:46:10 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

