

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
 Data File : VN071685.D
 Acq On : 31 Mar 2022 02:32
 Operator : JC\MD
 Sample : N2075-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 42 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RE137-20220322

Quant Time: Mar 31 07:46:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	649553	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1107644	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1106390	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	403525	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	368827	49.738	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.480%
35) Dibromofluoromethane	8.016	113	317533	48.717	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.440%
50) Toluene-d8	10.439	98	1360048	51.009	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.020%
62) 4-Bromofluorobenzene	12.727	95	461811	53.554	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.100%
Target Compounds						
					Qvalue	
9) 1,1,2-Trichlorotrifluo...	4.210	101	115255	20.457	ug/l	97
12) 1,1-Dichloroethene	4.187	96	31433	5.863	ug/l	97
16) Acetone	4.293	43	9736	3.072	ug/l	93
24) 1,1-Dichloroethane	6.381	63	10758	0.967	ug/l #	91
27) cis-1,2-Dichloroethene	7.322	96	22074	3.058	ug/l	89
29) Tetrahydrofuran	7.692	42	2820	0.953	ug/l #	84
30) Chloroform	7.816	83	14249	1.213	ug/l	98
38) Carbon Tetrachloride	8.198	117	19651	2.203	ug/l #	92
44) Trichloroethene	9.216	130	11836108	1597.852	ug/l	93
49) 1,4-Dioxane	9.575	88	2493	1.493	ug/l #	77
55) 1,1,2-Trichloroethane	10.892	97	6168	0.825	ug/l	94
64) Tetrachloroethene	10.974	164	21856	2.992	ug/l	95

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
Data File : VN071685.D
Acq On : 31 Mar 2022 02:32
Operator : JC\MD
Sample : N2075-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 42 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RE137-20220322

Quant Time: Mar 31 07:46:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

