

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
 Data File : VN071690.D
 Acq On : 31 Mar 2022 04:32
 Operator : JC\MD
 Sample : PB143727ZHE#11
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 47 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143727ZHE#11

Quant Time: Mar 31 07:47:39 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	618382	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1050345	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1065027	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	380540	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	357329	50.616	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 101.240%			
35) Dibromofluoromethane	8.022	113	301476	48.777	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 97.560%			
50) Toluene-d8	10.439	98	1302495	51.515	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 103.040%			
62) 4-Bromofluorobenzene	12.727	95	440021	53.811	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 107.620%			
Target Compounds						
16) Acetone	4.298	43	4959	1.644	ug/l	93
20) Methylene Chloride	5.104	84	6323	0.994	ug/l #	93
44) Trichloroethene	9.216	130	2677	0.381	ug/l	75

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
Data File : VN071690.D
Acq On : 31 Mar 2022 04:32
Operator : JC\MD
Sample : PB143727ZHE#11
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 47 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143727ZHE#11

Quant Time: Mar 31 07:47:39 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

