

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
 Data File : VN071695.D
 Acq On : 31 Mar 2022 06:33
 Operator : JC\MD
 Sample : N2167-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 52 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RAIL-PILE-1

Quant Time: Mar 31 07:48:43 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.081	168	651813	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1114016	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1158821	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.669	152	465708	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	375173	50.418	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.840%	
35) Dibromofluoromethane	8.016	113	319993	48.814	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.620%	
50) Toluene-d8	10.439	98	1419049	52.917	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.840%	
62) 4-Bromofluorobenzene	12.727	95	502439	57.932	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 115.860%	
Target Compounds						
					Qvalue	
16) Acetone	4.287	43	72087	22.668	ug/l	93
20) Methylene Chloride	5.093	84	493165	73.575	ug/l	99
43) Isopropyl Acetate	8.569	43	152112	9.127	ug/l #	69
95) Naphthalene	15.504	128	531839	26.974	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
Data File : VN071695.D
Acq On : 31 Mar 2022 06:33
Operator : JC\MD
Sample : N2167-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 52 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RAIL-PILE-1

Quant Time: Mar 31 07:48:43 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

