

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
 Data File : VN071664.D
 Acq On : 30 Mar 2022 17:20
 Operator : JC\MD
 Sample : N2130-03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 STORAGE-BLANK-REF-4

Quant Time: Mar 31 01:35:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 30 19:18:20 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	620109	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	1042502	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1027755	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.668	152	369077	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	358140	50.590	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.180%
35) Dibromofluoromethane	8.016	113	297584	48.510	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	97.020%
50) Toluene-d8	10.439	98	1279163	50.973	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	101.940%
62) 4-Bromofluorobenzene	12.727	95	429502	52.920	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	105.840%


Target Compounds	Qvalue


(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033022\
Data File : VN071664.D
Acq On : 30 Mar 2022 17:20
Operator : JC\MD
Sample : N2130-03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-REF-4

Quant Time: Mar 31 01:35:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 30 19:18:20 2022
Response via : Initial Calibration





Pentafluorobenzene

Concen: 50.000 ug/l

RT: 8.086 min Scan# 1036

Delta R.T. 0.005 min

Lab File: VN071664.D

Acq: 30 Mar 2022 17:20

Instrument :

MSVOA_N

ClientSampleId :





STORAGE-BLANK-REF-4

Tgt Ion:168 Resp: 620109

Ion Ratio Lower Upper

168 100

99 53.4 42.3 63.5



1,2-Dichloroethane-d4

Concen: 50.590 ug/l

RT: 8.433 min Scan# 1096

Delta R.T. -0.001 min

Lab File: VN071664.D




Acq: 30 Mar 2022 17:20


Tgt Ion: 65 Resp: 358140

Ion Ratio Lower Upper

65 100

67 51.9 0.0 104.8





#34

1,4-Difluorobenzene

Concen: 50.000 ug/l

RT: 8.963 min Scan# 1186

Delta R.T. -0.000 min

Lab File: VN071664.D

Acq: 30 Mar 2022 17:20

Instrument :

MSVOA_N

ClientSampleId :

STORAGE-BLANK-REF-4





Tgt Ion:114 Resp: 1042502

Ion Ratio Lower Upper

114 100

63 19.1 0.0 40.0

88 15.8 0.0 32.6



#35

Dibromofluoromethane

Concen: 48.510 ug/l

RT: 8.016 min Scan# 1025

Delta R.T. -0.000 min

Lab File: VN071664.D

Acq: 30 Mar 2022 17:20




Tgt Ion:113 Resp: 297584


Ion Ratio Lower Upper

113 100

111 103.1 81.4 122.0

192 22.1 17.0 25.6









#50
Toluene-d8
Concen: 50.973 ug/l
RT: 10.439 min Scan# 1437
Delta R.T. 0.000 min
Lab File: VN071664.D
Acq: 30 Mar 2022 17:20




Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-REF-4


Tgt Ion: 98 Resp: 1279163
Ion Ratio Lower Upper
98 100
100 64.5 52.6 78.8



#62
4-Bromofluorobenzene
Concen: 52.920 ug/l
RT: 12.727 min Scan# 1826
Delta R.T. 0.000 min
Lab File: VN071664.D
Acq: 30 Mar 2022 17:20

Tgt Ion: 95 Resp: 429502
Ion Ratio Lower Upper
95 100
174 84.7 0.0 167.6
176 81.8 0.0 163.4









#63
Chlorobenzene-d5
Concen: 50.000 ug/l
RT: 11.739 min Scan# 117
Delta R.T. 0.000 min
Lab File: VN071664.D
Acq: 30 Mar 2022 17:20

Instrument :
MSVOA_N
ClientSampleId :
STORAGE-BLANK-REF-4

Tgt Ion:117 Resp: 1027755
Ion Ratio Lower Upper
117 100
82 55.1 44.2 66.2
119 32.0 25.4 38.0



#72
1,4-Dichlorobenzene-d4
Concen: 50.000 ug/l
RT: 13.668 min Scan# 1986
Delta R.T. -0.001 min
Lab File: VN071664.D
Acq: 30 Mar 2022 17:20

Tgt Ion:152 Resp: 369077
Ion Ratio Lower Upper
152 100
115 56.3 28.9 86.7
150 157.7 0.0 356.2

