

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN033119\
 Data File : VN054849.D
 Acq On : 31 Mar 2019 14:39
 Operator : JC/SP
 Sample : K2179-08
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-4D

Quant Time: Apr 01 03:21:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	518795	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	866345	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	742488	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	263531	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	359385	51.39	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.78%	
35) Dibromofluoromethane	7.59	113	280227	48.52	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.04%	
50) Toluene-d8	10.09	98	1040857	48.47	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.94%	
62) 4-Bromofluorobenzene	12.40	95	320944	44.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	89.88%	
Target Compounds						
20) Methylene Chloride	4.54	84	371925	63.224	ug/l #	81
43) Isopropyl Acetate	8.17	43	19240	2.000	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN033119\
Data File : VN054849.D
Acq On : 31 Mar 2019 14:39
Operator : JC/SP
Sample : K2179-08
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-4D

Quant Time: Apr 01 03:21:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

