

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN033119\
 Data File : VN054850.D
 Acq On : 31 Mar 2019 15:06
 Operator : JC/SP
 Sample : K2179-12
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 11 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-8S

Quant Time: Apr 01 03:23:50 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	522348	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	871381	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	733450	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.35	152	253457	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	357891	50.83	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.66%	
35) Dibromofluoromethane	7.59	113	274398	47.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.46%	
50) Toluene-d8	10.09	98	1027326	47.57	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.14%	
62) 4-Bromofluorobenzene	12.40	95	312024	43.44	ug/l	0.00
Spiked Amount 50.000			Recovery	=	86.88%	
Target Compounds						
16) Acetone	3.83	43	11386	6.781	ug/l #	83
20) Methylene Chloride	4.54	84	38415	6.486	ug/l #	83
43) Isopropyl Acetate	8.17	43	19177	1.982	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN033119\
Data File : VN054850.D
Acq On : 31 Mar 2019 15:06
Operator : JC/SP
Sample : K2179-12
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 11 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-8S

Quant Time: Apr 01 03:23:50 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

