

Data Path : Z:\VOASRV\HPCHEM1\MSVOA_N\DATA\VN033119\
 Data File : VN054859.D
 Acq On : 31 Mar 2019 19:04
 Operator : JC/SP
 Sample : K2180-24
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-10D

Quant Time: Apr 01 04:23:24 2019
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
 Quant Title : SW846 8260
 QLast Update : Sat Mar 30 00:48:18 2019
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.67	168	544383	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.59	114	895420	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.41	117	747879	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	277187	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.03	65	364464	49.67	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.34%	
35) Dibromofluoromethane	7.59	113	281499	47.15	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.30%	
50) Toluene-d8	10.09	98	1059930	47.76	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.52%	
62) 4-Bromofluorobenzene	12.40	95	325161	44.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	88.10%	
Target Compounds						
20) Methylene Chloride	4.55	84	43747	7.087	ug/l #	83
43) Isopropyl Acetate	8.17	43	17969	1.807	ug/l #	95
67) Ethyl Benzene	11.51	91	29324	1.110	ug/l	99
68) m/p-Xylenes	11.62	106	24525	2.508	ug/l	97
80) 1,3,5-Trimethylbenzene	12.73	105	19364	1.042	ug/l	98
84) 1,2,4-Trimethylbenzene	13.04	105	50463	2.849	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA_N\DATA\VN033119\
Data File : VN054859.D
Acq On : 31 Mar 2019 19:04
Operator : JC/SP
Sample : K2180-24
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-10D

Quant Time: Apr 01 04:23:24 2019
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N032919W.M
Quant Title : SW846 8260
QLast Update : Sat Mar 30 00:48:18 2019
Response via : Initial Calibration

