

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033120\
 Data File : VN060781.D
 Acq On : 31 Mar 2020 13:26
 Operator : JC/MD
 Sample : PB127899TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 6 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127899TB

Quant Time: Apr 01 08:51:45 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	177827	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	293078	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	268390	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	113239	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	113410	46.46	ug/l	0.00
Spiked Amount 50.000			Recovery	=	92.92%	
35) Dibromofluoromethane	7.56	113	86092	48.58	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.16%	
50) Toluene-d8	10.08	98	356760	49.29	ug/l	0.00
Spiked Amount 50.000			Recovery	=	98.58%	
62) 4-Bromofluorobenzene	12.40	95	128954	48.61	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.22%	
Target Compounds						
16) Acetone	3.78	43	5236	6.055	ug/l	96
20) Methylene Chloride	4.52	84	4390	2.064	ug/l	96

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033120\
Data File : VN060781.D
Acq On : 31 Mar 2020 13:26
Operator : JC/MD
Sample : PB127899TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 6 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127899TB

Quant Time: Apr 01 08:51:45 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

