

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033120\
 Data File : VN060782.D
 Acq On : 31 Mar 2020 13:49
 Operator : JC/MD
 Sample : PB127897TB
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 7 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127897TB

Quant Time: Apr 01 08:56:02 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	198225	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	335521	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	314442	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	141637	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	130254	47.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	95.74%	
35) Dibromofluoromethane	7.56	113	98936	48.77	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.54%	
50) Toluene-d8	10.08	98	413837	49.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.88%	
62) 4-Bromofluorobenzene	12.40	95	154892	51.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	102.00%	
Target Compounds						
16) Acetone	3.79	43	6264	6.499	ug/l	93
20) Methylene Chloride	4.52	84	2920	1.231	ug/l	90
43) Isopropyl Acetate	8.14	43	67407	11.643	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033120\
Data File : VN060782.D
Acq On : 31 Mar 2020 13:49
Operator : JC/MD
Sample : PB127897TB
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 7 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127897TB

Quant Time: Apr 01 08:56:02 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

