

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033120\
 Data File : VN060785.D
 Acq On : 31 Mar 2020 14:57
 Operator : JC/MD
 Sample : L1758-06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 10 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PL-03-033020

Quant Time: Apr 01 09:03:58 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	209545	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	349867	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	328508	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	147801	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	135341	47.05	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.10%	
35) Dibromofluoromethane	7.56	113	103597	48.97	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.94%	
50) Toluene-d8	10.08	98	430964	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
62) 4-Bromofluorobenzene	12.40	95	161115	50.88	ug/l	0.00
Spiked Amount 50.000			Recovery	=	101.76%	
Target Compounds						
16) Acetone	3.78	43	11473	11.260	ug/l	96
20) Methylene Chloride	4.52	84	13492	5.383	ug/l	93
43) Isopropyl Acetate	8.14	43	62524	10.357	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033120\
Data File : VN060785.D
Acq On : 31 Mar 2020 14:57
Operator : JC/MD
Sample : L1758-06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 10 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PL-03-033020

Quant Time: Apr 01 09:03:58 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

