

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033120\
 Data File : VN060790.D
 Acq On : 31 Mar 2020 16:51
 Operator : JC/MD
 Sample : L2070-03
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 CHASSIS-WASH

Quant Time: Apr 01 09:19:27 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	201587	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	336983	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	316154	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	146387	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	130705	47.23	ug/l	0.00
Spiked Amount 50.000			Recovery	=	94.46%	
35) Dibromofluoromethane	7.56	113	99316	48.74	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.48%	
50) Toluene-d8	10.08	98	415817	49.96	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.92%	
62) 4-Bromofluorobenzene	12.40	95	158398	51.93	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.86%	
Target Compounds						
16) Acetone	3.78	43	67486	68.846	ug/l	98
20) Methylene Chloride	4.52	84	5523	2.290	ug/l	97
25) 2-Butanone	6.80	43	40940	25.810	ug/l	97
43) Isopropyl Acetate	8.14	43	33468	5.756	ug/l #	91
51) 4-Methyl-2-Pentanone	9.97	43	212924	64.350	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033120\
Data File : VN060790.D
Acq On : 31 Mar 2020 16:51
Operator : JC/MD
Sample : L2070-03
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
CHASSIS-WASH

Quant Time: Apr 01 09:19:27 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

