

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033120\
 Data File : VN060794.D
 Acq On : 31 Mar 2020 18:21
 Operator : JC/MD
 Sample : PB127897ZHE#01
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127897ZHE#01

Quant Time: Apr 01 09:08:23 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	202672	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.56	114	337757	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	319265	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	149368	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	131898	47.41	ug/l	0.00
Spiked Amount	50.000		Recovery	=	94.82%	
35) Dibromofluoromethane	7.56	113	100643	49.28	ug/l	0.00
Spiked Amount	50.000		Recovery	=	98.56%	
50) Toluene-d8	10.08	98	422476	50.64	ug/l	0.00
Spiked Amount	50.000		Recovery	=	101.28%	
62) 4-Bromofluorobenzene	12.40	95	159288	52.10	ug/l	0.00
Spiked Amount	50.000		Recovery	=	104.20%	
Target Compounds						
16) Acetone	3.78	43	6799	6.899	ug/l	97
43) Isopropyl Acetate	8.14	43	62775	10.771	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033120\
Data File : VN060794.D
Acq On : 31 Mar 2020 18:21
Operator : JC/MD
Sample : PB127897ZHE#01
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127897ZHE#01

Quant Time: Apr 01 09:08:23 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

