

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033120\
 Data File : VN060795.D
 Acq On : 31 Mar 2020 18:44
 Operator : JC/MD
 Sample : PB127897ZHE#02
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127897ZHE#02

Quant Time: Apr 01 09:05:42 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	209902	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	348950	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	332453	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	153495	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	135122	46.89	ug/l	0.00
Spiked Amount 50.000			Recovery	=	93.78%	
35) Dibromofluoromethane	7.56	113	103276	48.95	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.90%	
50) Toluene-d8	10.08	98	434536	50.42	ug/l	0.00
Spiked Amount 50.000			Recovery	=	100.84%	
62) 4-Bromofluorobenzene	12.40	95	164071	51.94	ug/l	0.00
Spiked Amount 50.000			Recovery	=	103.88%	
Target Compounds						
16) Acetone	3.79	43	6132	6.008	ug/l	99
43) Isopropyl Acetate	8.14	43	59908	9.950	ug/l #	90

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033120\
Data File : VN060795.D
Acq On : 31 Mar 2020 18:44
Operator : JC/MD
Sample : PB127897ZHE#02
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127897ZHE#02

Quant Time: Apr 01 09:05:42 2020
Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

