

Data Path : Z:\VOASRV\HPCHEM1\MSVOA N\DATA\VN033120\
 Data File : VN060799.D
 Acq On : 31 Mar 2020 20:15
 Operator : JC/MD
 Sample : PB127897ZHE#06
 Misc : 5.00mL/MSVOA N/WATER
 ALS Vial : 24 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB127897ZHE#06

Quant Time: Apr 01 08:56:37 2020
 Quant Method : Z:\VOASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 18 08:49:09 2020
 Response via : Initial Calibration

Internal Standards	R.T.	QIon	Response	Conc	Units	Dev(Min)
1) Pentafluorobenzene	7.64	168	199696	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.57	114	336489	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.40	117	320464	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.34	152	148267	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.00	65	131582	48.00	ug/l	0.00
Spiked Amount 50.000			Recovery	=	96.00%	
35) Dibromofluoromethane	7.56	113	98890	48.60	ug/l	0.00
Spiked Amount 50.000			Recovery	=	97.20%	
50) Toluene-d8	10.08	98	414440	49.87	ug/l	0.00
Spiked Amount 50.000			Recovery	=	99.74%	
62) 4-Bromofluorobenzene	12.40	95	159639	52.41	ug/l	0.00
Spiked Amount 50.000			Recovery	=	104.82%	
Target Compounds						
16) Acetone	3.78	43	6271	6.458	ug/l	92
20) Methylene Chloride	4.53	84	3189	1.335	ug/l	84
43) Isopropyl Acetate	8.14	43	64474	11.104	ug/l #	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\V0ASRV\HPCHEM1\MSVOA N\DATA\VN033120\
Data File : VN060799.D
Acq On : 31 Mar 2020 20:15
Operator : JC/MD
Sample : PB127897ZHE#06
Misc : 5.00mL/MSVOA N/WATER
ALS Vial : 24 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB127897ZHE#06

Quant Time: Apr 01 08:56:37 2020
Quant Method : Z:\V0ASRV\HPCHEM1\MSVOA_N\METHODS\82N031820W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 18 08:49:09 2020
Response via : Initial Calibration

