

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033121\
 Data File : VN066441.D
 Acq On : 31 Mar 2021 14:11
 Operator : JC/MD
 Sample : M1838-02
 Misc : 5.00mL/MSVOA_N/WATER
 ALS Vial : 8 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PHC-BEAPT-001

Manual Integrations
 APPROVED

MMDadoda
 4/1/2021 4:44:48 PM

Quant Time: Apr 01 01:37:13 2021
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
 Quant Title : SW846 8260
 QLast Update : Fri Mar 26 16:26:01 2021
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	7.585	168	443884	50.00	ug/l	0.00
34) 1,4-Difluorobenzene	8.515	114	673440	50.00	ug/l	0.00
63) Chlorobenzene-d5	11.347	117	613971	50.00	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.289	152	217559	50.00	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	7.947	65	288626	48.91	ug/l	0.00
Spiked Amount 50.000	Range 78 - 117		Recovery =	97.82%		
35) Dibromofluoromethane	7.507	113	219807	49.24	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery =	98.48%		
50) Toluene-d8	10.028	98	898753	53.08	ug/l	0.00
Spiked Amount 50.000	Range 92 - 112		Recovery =	106.16%		
62) 4-Bromofluorobenzene	12.345	95	252069	42.86	ug/l	0.00
Spiked Amount 50.000	Range 83 - 123		Recovery =	85.72%		
Target Compounds						
						Qvalue
16) Acetone	3.739	43	131030	41.10	ug/l	97
18) Methyl Acetate	4.237	43	135415	16.26	ug/l	99
20) Methylene Chloride	4.452	84	38342	5.65	ug/l	97
25) 2-Butanone	6.756	43	17858	5.12	ug/l	97
43) Isopropyl Acetate	8.091	43	32240	2.71	ug/l	98
95) Naphthalene	15.081	128	10319m	1.87	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033121\
Data File : VN066441.D
Acq On : 31 Mar 2021 14:11
Operator : JC/MD
Sample : M1838-02
Misc : 5.00mL/MSVOA_N/WATER
ALS Vial : 8 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PHC-BEAPT-001

Manual Integrations
APPROVED

MMDadoda
4/1/2021 4:44:48 PM

Quant Time: Apr 01 01:37:13 2021
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N032521W.M
Quant Title : SW846 8260
QLast Update : Fri Mar 26 16:26:01 2021
Response via : Initial Calibration

