

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
 Data File : VN071711.D
 Acq On : 31 Mar 2022 18:01
 Operator : JC\MD
 Sample : N2075-13DL 10X
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 14 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW17811-20220322DL

Quant Time: Apr 01 01:38:53 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	591033	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.963	114	991461	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	985983	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	338096	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	337928	50.083	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.160%	
35) Dibromofluoromethane	8.022	113	282601	48.439	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 96.880%	
50) Toluene-d8	10.439	98	1239384	51.930	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.860%	
62) 4-Bromofluorobenzene	12.727	95	391215	50.684	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 101.360%	
Target Compounds						Qvalue
27) cis-1,2-Dichloroethene	7.328	96	31485	4.794	ug/l	88
44) Trichloroethene	9.216	130	334019	50.376	ug/l	100
45) 1,2-Dichloropropane	9.486	63	16582	2.579	ug/l #	88
64) Tetrachloroethene	10.975	164	18491	2.840	ug/l	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
Data File : VN071711.D
Acq On : 31 Mar 2022 18:01
Operator : JC\MD
Sample : N2075-13DL 10X
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 14 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW17811-20220322DL

Quant Time: Apr 01 01:38:53 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

