

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
 Data File : VN071721.D
 Acq On : 31 Mar 2022 22:50
 Operator : JC\MD
 Sample : VN0331WBL03
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 26 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0331WBL03

Quant Time: Apr 01 01:41:27 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	672410	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1138442	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1155160	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	414172	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.439	65	383777	49.995	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.980%
35) Dibromofluoromethane	8.022	113	322743	48.177	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.360%
50) Toluene-d8	10.439	98	1416789	51.699	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.400%
62) 4-Bromofluorobenzene	12.727	95	474803	53.571	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	107.140%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
Data File : VN071721.D
Acq On : 31 Mar 2022 22:50
Operator : JC\MD
Sample : VN0331WBL03
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 26 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0331WBL03

Quant Time: Apr 01 01:41:27 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

