

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
 Data File : VN071725.D
 Acq On : 01 Apr 2022 00:26
 Operator : JC\MD
 Sample : N2090-22
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 30 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB032122

Quant Time: Apr 01 01:42:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	670320	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1141664	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1154475	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	415093	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	387668	50.659	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.320%
35) Dibromofluoromethane	8.022	113	329282	49.015	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	98.020%
50) Toluene-d8	10.439	98	1432444	52.123	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	104.240%
62) 4-Bromofluorobenzene	12.727	95	464755	52.290	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.580%
Target Compounds						
16) Acetone	4.293	43	6123	1.872	ug/l	91

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
Data File : VN071725.D
Acq On : 01 Apr 2022 00:26
Operator : JC\MD
Sample : N2090-22
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 30 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB032122

Quant Time: Apr 01 01:42:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

