

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
 Data File : VN071727.D
 Acq On : 01 Apr 2022 01:14
 Operator : JC\MD
 Sample : N2090-23
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 32 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB032222

Quant Time: Apr 01 06:03:52 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	644715	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1099193	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1088533	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	382059	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	371877	50.526	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	101.060%
35) Dibromofluoromethane	8.022	113	312680	48.342	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.680%
50) Toluene-d8	10.439	98	1362882	51.508	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	103.020%
62) 4-Bromofluorobenzene	12.727	95	444643	51.960	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	103.920%
Target Compounds						
						Qvalue
16) Acetone	4.293	43	5955	1.893	ug/l	95
20) Methylene Chloride	5.098	84	2914	0.440	ug/l #	93

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
Data File : VN071727.D
Acq On : 01 Apr 2022 01:14
Operator : JC\MD
Sample : N2090-23
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 32 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB032222

Quant Time: Apr 01 06:03:52 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

