

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
 Data File : VN071730.D
 Acq On : 01 Apr 2022 02:26
 Operator : JC\MD
 Sample : N2200-19
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 35 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 FB032922

Quant Time: Apr 01 06:04:41 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	657103	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1104912	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1133102	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	404588	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	377617	50.338	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.680%	
35) Dibromofluoromethane	8.022	113	320195	49.247	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 98.500%	
50) Toluene-d8	10.439	98	1403287	52.761	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 105.520%	
62) 4-Bromofluorobenzene	12.727	95	460430	53.526	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 107.060%	
Target Compounds						
16) Acetone	4.298	43	5580	1.741	ug/l	Qvalue 89

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
Data File : VN071730.D
Acq On : 01 Apr 2022 02:26
Operator : JC\MD
Sample : N2200-19
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 35 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
FB032922

Quant Time: Apr 01 06:04:41 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

