

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
 Data File : VN071732.D
 Acq On : 01 Apr 2022 03:15
 Operator : JC\MD
 Sample : N2090-06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 37 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-50D

Quant Time: Apr 01 06:05:12 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	660804	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1113813	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1106677	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	394173	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.439	65	373417	49.500	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	=	99.000%
35) Dibromofluoromethane	8.022	113	317792	48.487	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	=	96.980%
50) Toluene-d8	10.439	98	1378357	51.409	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	=	102.820%
62) 4-Bromofluorobenzene	12.727	95	455197	52.495	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	=	104.980%
Target Compounds						
					Qvalue	
11) Tert butyl alcohol	5.369	59	9676	7.904	ug/l #	72
16) Acetone	4.298	43	8713	2.703	ug/l	98
19) Methyl tert-butyl Ether	5.622	73	226603	11.183	ug/l	96
22) Diisopropyl ether	6.498	45	8525	0.442	ug/l #	79

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
Data File : VN071732.D
Acq On : 01 Apr 2022 03:15
Operator : JC\MD
Sample : N2090-06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 37 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-50D

Quant Time: Apr 01 06:05:12 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

