

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
 Data File : VN071735.D
 Acq On : 01 Apr 2022 04:27
 Operator : JC\MD
 Sample : N2090-18
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 40 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MW-62

Quant Time: Apr 01 06:06:08 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	675538	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1156506	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1155222	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	416167	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.439	65	387639	50.264	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 100.520%	
35) Dibromofluoromethane	8.022	113	330334	48.540	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 97.080%	
50) Toluene-d8	10.439	98	1443461	51.850	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 103.700%	
62) 4-Bromofluorobenzene	12.727	95	463464	51.475	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 102.960%	
Target Compounds						
					Qvalue	
16) Acetone	4.299	43	8323	2.525	ug/l #	82
39) Methylcyclohexane	9.457	83	3707	0.328	ug/l #	88
67) Ethyl Benzene	11.839	91	12163	0.311	ug/l	97
73) Isopropylbenzene	12.574	105	37050	1.056	ug/l	100
78) n-propylbenzene	12.916	91	54057	1.465	ug/l	99
85) sec-Butylbenzene	13.498	105	15441	0.469	ug/l	80

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
Data File : VN071735.D
Acq On : 01 Apr 2022 04:27
Operator : JC\MD
Sample : N2090-18
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 40 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MW-62

Quant Time: Apr 01 06:06:08 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

