

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
 Data File : VN071738.D
 Acq On : 01 Apr 2022 05:40
 Operator : JC\MD
 Sample : PB143762ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 43 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB143762ZHE#02

Quant Time: Apr 01 06:07:00 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.087	168	631703	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1099923	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.739	117	1135214	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	412635	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.434	65	364258	50.510	ug/l	0.00
Spiked Amount	50.000	Range	61 - 141	Recovery	= 101.020%	
35) Dibromofluoromethane	8.022	113	312121	48.223	ug/l	0.00
Spiked Amount	50.000	Range	69 - 133	Recovery	= 96.440%	
50) Toluene-d8	10.439	98	1413947	53.403	ug/l	0.00
Spiked Amount	50.000	Range	65 - 126	Recovery	= 106.800%	
62) 4-Bromofluorobenzene	12.727	95	474404	55.401	ug/l	0.00
Spiked Amount	50.000	Range	58 - 135	Recovery	= 110.800%	
Target Compounds						Qvalue
16) Acetone	4.293	43	45130	14.643	ug/l	96
20) Methylene Chloride	5.098	84	543094	83.604	ug/l	99
43) Isopropyl Acetate	8.581	43	151230	9.191	ug/l #	53

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
Data File : VN071738.D
Acq On : 01 Apr 2022 05:40
Operator : JC\MD
Sample : PB143762ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 43 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB143762ZHE#02

Quant Time: Apr 01 06:07:00 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

