

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
 Data File : VN071739.D
 Acq On : 01 Apr 2022 06:05
 Operator : JC\MD
 Sample : N2202-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 44 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 MHE

Manual Integrations
 APPROVED

Reviewed By : John Carlone 04/01/2022
 Supervised By : Mahesh Dadoda 04/01/2022

Quant Time: Apr 01 07:47:06 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.086	168	649956	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	8.969	114	1117004	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.745	117	1147835	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.674	152	409129	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.433	65	373735	50.369	ug/l	0.00
Spiked Amount 50.000	Range 61 - 141		Recovery = 100.740%			
35) Dibromofluoromethane	8.022	113	318493	48.455	ug/l	0.00
Spiked Amount 50.000	Range 69 - 133		Recovery = 96.920%			
50) Toluene-d8	10.439	98	1424389	52.974	ug/l	0.00
Spiked Amount 50.000	Range 65 - 126		Recovery = 105.940%			
62) 4-Bromofluorobenzene	12.727	95	465483	53.528	ug/l	0.00
Spiked Amount 50.000	Range 58 - 135		Recovery = 107.060%			
Target Compounds					Qvalue	
16) Acetone	4.293	43	53771	16.957	ug/l	97
20) Methylene Chloride	5.098	84	476254	71.255	ug/l	99
43) Isopropyl Acetate	8.580	43	146286m	8.754	ug/l	

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
Data File : VN071739.D
Acq On : 01 Apr 2022 06:05
Operator : JC\MD
Sample : N2202-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 44 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
MHE

Quant Time: Apr 01 07:47:06 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

Manual Integrations
APPROVED

Reviewed By :John Carlone 04/01/2022
Supervised By :Mahesh Dadoda 04/01/2022

