

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
 Data File : VN071718.D
 Acq On : 31 Mar 2022 20:50
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 01 01:40:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
1 I	Pentafluorobenzene	1.000	1.000	0.0	101	0.00
2 T	Dichlorodifluoromethane	0.398	0.391	1.8	95	0.00
3 P	Chloromethane	0.504	0.466	7.5	98	0.00
4 C	Vinyl Chloride	0.553	0.534	3.4#	95	0.00
5 T	Bromomethane	0.360	0.327	9.2	90	0.00
6 T	Chloroethane	0.398	0.373	6.3	94	0.00
7 T	Trichlorofluoromethane	0.704	0.676	4.0	95	0.00
8 T	Diethyl Ether	0.297	0.298	-0.3	98	0.00
9 T	1,1,2-Trichlorotrifluoroeth	0.434	0.417	3.9	96	0.00
10 T	Methyl Iodide	0.606	0.603	0.5	96	0.00
11 T	Tert butyl alcohol	0.093	0.101	-8.6	101	0.00
12 CM	1,1-Dichloroethene	0.413	0.399	3.4#	96	0.00
13 T	Acrolein	0.113	0.122	-8.0	101	0.00
14 T	Allyl chloride	0.670	0.668	0.3	97	0.00
15 T	Acrylonitrile	0.269	0.294	-9.3	103	0.00
16 T	Acetone	0.244	0.225	7.8	97	0.01
17 T	Carbon Disulfide	0.947	0.867	8.4	95	0.00
18 T	Methyl Acetate	0.737	0.630	14.5	101	0.00
19 T	Methyl tert-butyl Ether	1.533	1.588	-3.6	99	0.00
20 T	Methylene Chloride	0.514	0.505	1.8	101	0.00
21 T	trans-1,2-Dichloroethene	0.453	0.434	4.2	97	0.00
22 T	Diisopropyl ether	1.458	1.501	-2.9	100	0.00
23 T	Vinyl Acetate	1.160	1.276	-10.0	98	0.00
24 P	1,1-Dichloroethane	0.856	0.867	-1.3	99	0.00
25 T	2-Butanone	0.352	0.368	-4.5	101	0.00
26 T	2,2-Dichloropropane	0.735	0.667	9.3	88	0.00
27 T	cis-1,2-Dichloroethene	0.556	0.560	-0.7	99	0.00
28 T	Bromochloromethane	0.345	0.358	-3.8	114	0.00
29 T	Tetrahydrofuran	0.228	0.244	-7.0	100	0.00
30 C	Chloroform	0.904	0.898	0.7#	98	0.00
31 T	Cyclohexane	0.776	0.714	8.0	96	0.00
32 T	1,1,1-Trichloroethane	0.773	0.769	0.5	97	0.00
33 S	1,2-Dichloroethane-d4	0.571	0.593	-3.9	105	0.00
34 I	1,4-Difluorobenzene	1.000	1.000	0.0	103	0.00
35 S	Dibromofluoromethane	0.294	0.304	-3.4	106	0.00
36 T	1,1-Dichloropropene	0.411	0.389	5.4	97	0.00
37 T	Ethyl Acetate	0.450	0.456	-1.3	97	0.00
38 T	Carbon Tetrachloride	0.403	0.393	2.5	97	0.00
39 T	Methylcyclohexane	0.489	0.472	3.5	95	0.00
40 TM	Benzene	1.298	1.256	3.2	100	0.00
41 T	Methacrylonitrile	0.220	0.237	-7.7	106	0.00
42 TM	1,2-Dichloroethane	0.449	0.424	5.6	97	0.00
43 T	Isopropyl Acetate	0.748	0.708	5.3	100	0.00
44 TM	Trichloroethene	0.334	0.322	3.6	97	0.00
45 C	1,2-Dichloropropane	0.324	0.321	0.9#	101	0.00
46 T	Dibromomethane	0.220	0.214	2.7	99	0.00
47 T	Bromodichloromethane	0.433	0.440	-1.6	99	0.00
48 T	Methyl methacrylate	0.310	0.338	-9.0	106	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
 Data File : VN071718.D
 Acq On : 31 Mar 2022 20:50
 Operator : JC\MD
 Sample : VSTDCCC050
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 21 Sample Multiplier: 1

Instrument :
 MSVOA_N
 LabSampleId :
 VSTDCCC050

Quant Time: Apr 01 01:40:38 2022
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
 Quant Title : SW846 8260
 QLast Update : Thu Mar 31 01:57:54 2022
 Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
 Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
49 T	1,4-Dioxane	0.008	0.007	12.5	101	0.00
50 S	Toluene-d8	1.204	1.238	-2.8	106	0.00
51 T	4-Methyl-2-Pentanone	0.430	0.463	-7.7	101	0.00
52 CM	Toluene	0.820	0.817	0.4#	99	0.00
53 T	t-1,3-Dichloropropene	0.454	0.465	-2.4	97	0.00
54 T	cis-1,3-Dichloropropene	0.502	0.519	-3.4	99	0.00
55 T	1,1,2-Trichloroethane	0.338	0.334	1.2	100	0.00
56 T	Ethyl methacrylate	0.482	0.531	-10.2	102	0.00
57 T	1,3-Dichloropropane	0.565	0.560	0.9	100	0.00
58 T	2-Chloroethyl Vinyl ether	0.111	0.112	-0.9	83	0.00
59 T	2-Hexanone	0.303	0.327	-7.9	99	0.00
60 T	Dibromochloromethane	0.329	0.350	-6.4	101	0.00
61 T	1,2-Dibromoethane	0.332	0.337	-1.5	100	0.00
62 S	4-Bromofluorobenzene	0.389	0.396	-1.8	100	0.00
63 I	Chlorobenzene-d5	1.000	1.000	0.0	103	0.00
64 T	Tetrachloroethene	0.330	0.303	8.2	97	0.00
65 PM	Chlorobenzene	1.001	0.964	3.7	98	0.00
66 T	1,1,1,2-Tetrachloroethane	0.361	0.358	0.8	100	0.00
67 C	Ethyl Benzene	1.695	1.703	-0.5#	98	0.00
68 T	m/p-Xylenes	0.651	0.648	0.5	95	0.00
69 T	o-Xylene	0.646	0.648	-0.3	96	0.00
70 T	Styrene	0.979	1.032	-5.4	96	0.00
71 P	Bromoform	0.275	0.291	-5.8	98	0.00
72 I	1,4-Dichlorobenzene-d4	1.000	1.000	0.0	94	0.00
73 T	Isopropylbenzene	4.213	4.264	-1.2	96	0.00
74 T	N-amyl acetate	1.207	1.331	-10.3	93	0.00
75 P	1,1,2,2-Tetrachloroethane	1.406	1.375	2.2	96	0.00
76 T	1,2,3-Trichloropropane	1.158	1.051	9.2	88	0.00
77 T	Bromobenzene	1.056	1.006	4.7	93	0.00
78 T	n-propylbenzene	4.432	4.519	-2.0	93	0.00
79 T	2-Chlorotoluene	2.842	2.845	-0.1	96	0.00
80 T	1,3,5-Trimethylbenzene	3.338	3.410	-2.2	96	0.00
81 T	trans-1,4-Dichloro-2-butene	0.374	0.384	-2.7	89	0.00
82 T	4-Chlorotoluene	2.635	2.606	1.1	92	0.00
83 T	tert-Butylbenzene	3.052	3.083	-1.0	94	0.00
84 T	1,2,4-Trimethylbenzene	3.257	3.329	-2.2	95	0.00
85 T	sec-Butylbenzene	3.956	4.080	-3.1	93	0.00
86 T	p-Isopropyltoluene	3.154	3.312	-5.0	92	0.00
87 T	1,3-Dichlorobenzene	1.683	1.640	2.6	90	0.00
88 T	1,4-Dichlorobenzene	1.642	1.557	5.2	89	0.00
89 T	n-Butylbenzene	2.339	2.408	-2.9	89	0.00
90 T	Hexachloroethane	0.577	0.611	-5.9	95	0.00
91 T	1,2-Dichlorobenzene	1.652	1.588	3.9	91	0.00
92 T	1,2-Dibromo-3-Chloropropane	0.218	0.235	-7.8	95	0.00
93 T	1,2,4-Trichlorobenzene	0.649	0.750	-15.6	89	0.00
94 T	Hexachlorobutadiene	0.487	0.432	11.3	85	0.00
95 T	Naphthalene	1.586	2.117	-33.5#	91	0.00

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
Data File : VN071718.D
Acq On : 31 Mar 2022 20:50
Operator : JC\MD
Sample : VSTDCCC050
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 21 Sample Multiplier: 1

Instrument :
MSVOA_N
LabSampleId :
VSTDCCC050

Quant Time: Apr 01 01:40:38 2022
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Quant Title : SW846 8260
QLast Update : Thu Mar 31 01:57:54 2022
Response via : Initial Calibration

Min. RRF : 0.000 Min. Rel. Area : 50% Max. R.T. Dev 0.50min
Max. RRF Dev : 25% Max. Rel. Area : 150%

	Compound	AvgRF	CCRF	%Dev	Area%	Dev(min)
96 T	1,2,3-Trichlorobenzene	0.604	0.736	-21.9	90	0.00

(#) = Out of Range

SPCC's out = 0 CCC's out = 6