


Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033122\
Data File : VN071719.D
Acq On : 31 Mar 2022 21:38
Operator : JC\MD
Sample : BFB
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 23 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
BFB

Integration File: RTEINT.P

Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N033022W.M
Title : SW846 8260
Last Update : Thu Mar 31 01:57:54 2022



AutoFind: Scans 1825, 1826, 1827; Background Corrected with Scan 1817

Target	Rel. to	Lower	Upper	Rel.	Raw	Result
Mass	Mass	Limit%	Limit%	Abn%	Abn	Pass/Fail
50	95	15	40	16.8	13249	PASS
75	95	30	60	46.5	36589	PASS
95	95	100	100	100.0	78635	PASS
96	95	5	9	6.6	5167	PASS
173	174	0.00	2	0.7	457	PASS
174	95	50	100	83.3	65515	PASS
175	174	5	9	7.4	4824	PASS
176	174	95	101	95.7	62669	PASS
177	176	5	9	6.5	4073	PASS