

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
 Data File : VN086169.D
 Acq On : 31 Mar 2025 14:51
 Operator : JC\MD
 Sample : VN0331MBL01
 Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
 ALS Vial : 4 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 VN0331MBL01

Quant Time: Apr 01 01:36:11 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	175835	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	324396	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	288485	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	114979	50.000	ug/l	0.00

System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	129652	53.847	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	107.700%
35) Dibromofluoromethane	8.165	113	116836	51.600	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.200%
50) Toluene-d8	10.565	98	389121	47.352	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	94.700%
62) 4-Bromofluorobenzene	12.847	95	116915	39.894	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	79.780%

Target Compounds	Qvalue

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
Data File : VN086169.D
Acq On : 31 Mar 2025 14:51
Operator : JC\MD
Sample : VN0331MBL01
Misc : 5.00g/10mL/100uL/5.00mL/MSVOA_N/MEOH
ALS Vial : 4 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
VN0331MBL01

Quant Time: Apr 01 01:36:11 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

