

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
 Data File : VN086178.D
 Acq On : 31 Mar 2025 18:39
 Operator : JC\MD
 Sample : PB167374ZHE#02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 13 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167374ZHE#02

Quant Time: Apr 01 01:39:17 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	181948	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	343213	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	330122	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	123339	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	137306	55.110	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	110.220%	
35) Dibromofluoromethane	8.165	113	124054	51.784	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	103.560%	
50) Toluene-d8	10.565	98	447388	51.458	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	102.920%	
62) 4-Bromofluorobenzene	12.847	95	130386	42.051	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	84.100%	
Target Compounds						
						Qvalue
16) Acetone	4.430	43	42184	55.420	ug/l	99
20) Methylene Chloride	5.277	84	121332	54.488	ug/l	96
43) Isopropyl Acetate	8.688	43	29712	5.152	ug/l #	68

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
Data File : VN086178.D
Acq On : 31 Mar 2025 18:39
Operator : JC\MD
Sample : PB167374ZHE#02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 13 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167374ZHE#02

Quant Time: Apr 01 01:39:17 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

