

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
 Data File : VN086180.D
 Acq On : 31 Mar 2025 19:27
 Operator : JC\MD
 Sample : PB167374ZHE#04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 15 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167374ZHE#04

Quant Time: Apr 01 01:39:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168235	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	331623	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	319378	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	117711	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	131644	57.144	ug/l	0.00
Spiked Amount 50.000	Range 74 - 125		Recovery = 114.280%			
35) Dibromofluoromethane	8.165	113	119318	51.548	ug/l	0.00
Spiked Amount 50.000	Range 75 - 124		Recovery = 103.100%			
50) Toluene-d8	10.565	98	425361	50.634	ug/l	0.00
Spiked Amount 50.000	Range 86 - 113		Recovery = 101.260%			
62) 4-Bromofluorobenzene	12.847	95	123959	41.376	ug/l	0.00
Spiked Amount 50.000	Range 77 - 121		Recovery = 82.760%			
Target Compounds					Qvalue	
16) Acetone	4.430	43	40270	57.218	ug/l	97
20) Methylene Chloride	5.277	84	121036	58.786	ug/l	96
43) Isopropyl Acetate	8.688	43	40429	7.256	ug/l #	64

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
Data File : VN086180.D
Acq On : 31 Mar 2025 19:27
Operator : JC\MD
Sample : PB167374ZHE#04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 15 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167374ZHE#04

Quant Time: Apr 01 01:39:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

