

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
 Data File : VN086181.D
 Acq On : 31 Mar 2025 19:51
 Operator : JC\MD
 Sample : PB167374ZHE#05
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 16 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167374ZHE#05

Quant Time: Apr 01 01:40:08 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	178487	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	347453	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	328489	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	118630	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	138830	56.802	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	113.600%	
35) Dibromofluoromethane	8.165	113	123757	51.030	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	102.060%	
50) Toluene-d8	10.565	98	444676	50.522	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	101.040%	
62) 4-Bromofluorobenzene	12.847	95	124595	39.693	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	79.380%	
Target Compounds						
						Qvalue
16) Acetone	4.424	43	32789	43.912	ug/l	93
20) Methylene Chloride	5.277	84	89382	40.918	ug/l	100
43) Isopropyl Acetate	8.688	43	42220	7.232	ug/l #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
Data File : VN086181.D
Acq On : 31 Mar 2025 19:51
Operator : JC\MD
Sample : PB167374ZHE#05
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 16 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167374ZHE#05

Quant Time: Apr 01 01:40:08 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

