

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
 Data File : VN086182.D
 Acq On : 31 Mar 2025 20:16
 Operator : JC\MD
 Sample : PB167374ZHE#06
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 17 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 PB167374ZHE#06

Quant Time: Apr 01 01:40:22 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	173060	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	342725	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	325187	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	115982	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	135476	57.168	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	114.340%
35) Dibromofluoromethane	8.165	113	123285	51.537	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.080%
50) Toluene-d8	10.565	98	443618	51.097	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	102.200%
62) 4-Bromofluorobenzene	12.847	95	123370	39.845	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	79.700%
Target Compounds						
						Qvalue
16) Acetone	4.430	43	50511	69.767	ug/l	94
20) Methylene Chloride	5.277	84	123351	58.240	ug/l	96
43) Isopropyl Acetate	8.683	43	50080	8.697	ug/l #	62

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
Data File : VN086182.D
Acq On : 31 Mar 2025 20:16
Operator : JC\MD
Sample : PB167374ZHE#06
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 17 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
PB167374ZHE#06

Quant Time: Apr 01 01:40:22 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

