

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
 Data File : VN086184.D
 Acq On : 31 Mar 2025 21:04
 Operator : JC\MD
 Sample : Q1674-02
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 19 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 RT5358

Quant Time: Apr 01 05:23:52 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	168391	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	332486	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	301902	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	112961	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	132677	57.539	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	=	115.080%	
35) Dibromofluoromethane	8.165	113	120041	51.726	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	=	103.460%	
50) Toluene-d8	10.565	98	415940	49.384	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	=	98.760%	
62) 4-Bromofluorobenzene	12.847	95	122296	40.715	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	=	81.420%	
Target Compounds						
						Qvalue
16) Acetone	4.436	43	49146	69.764	ug/l	98
20) Methylene Chloride	5.277	84	106507	51.681	ug/l	98
43) Isopropyl Acetate	8.694	43	44265	7.924	ug/l #	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
Data File : VN086184.D
Acq On : 31 Mar 2025 21:04
Operator : JC\MD
Sample : Q1674-02
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 19 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
RT5358

Quant Time: Apr 01 05:23:52 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

