

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
 Data File : VN086185.D
 Acq On : 31 Mar 2025 21:28
 Operator : JC\MD
 Sample : Q1674-04
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 20 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 72-11991

Quant Time: Apr 01 05:29:47 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T.	QIon	Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	162466	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	327287	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	299534	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	105464	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	131716	59.206	ug/l	0.00
Spiked Amount	50.000	Range	74 - 125	Recovery	=	118.420%
35) Dibromofluoromethane	8.165	113	118459	51.855	ug/l	0.00
Spiked Amount	50.000	Range	75 - 124	Recovery	=	103.720%
50) Toluene-d8	10.565	98	410386	49.499	ug/l	0.00
Spiked Amount	50.000	Range	86 - 113	Recovery	=	99.000%
62) 4-Bromofluorobenzene	12.847	95	119451	40.399	ug/l	0.00
Spiked Amount	50.000	Range	77 - 121	Recovery	=	80.800%
Target Compounds						
					Qvalue	
16) Acetone	4.430	43	56049	82.465	ug/l	99
20) Methylene Chloride	5.277	84	104784	52.700	ug/l	97
25) 2-Butanone	7.482	43	5195	5.297	ug/l	97
43) Isopropyl Acetate	8.688	43	50261	9.140	ug/l #	58

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
Data File : VN086185.D
Acq On : 31 Mar 2025 21:28
Operator : JC\MD
Sample : Q1674-04
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 20 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
72-11991

Quant Time: Apr 01 05:29:47 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

