

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
 Data File : VN086187.D
 Acq On : 31 Mar 2025 22:17
 Operator : JC\MD
 Sample : Q1679-08
 Misc : 5.0mL/MSVOA_N/WATER
 ALS Vial : 22 Sample Multiplier: 1

Instrument :
 MSVOA_N
 ClientSampleId :
 TP-9

Quant Time: Apr 01 01:41:42 2025
 Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
 Quant Title : SW846 8260
 QLast Update : Wed Mar 19 03:20:56 2025
 Response via : Initial Calibration

Compound	R.T. QIon		Response	Conc	Units	Dev(Min)

Internal Standards						
1) Pentafluorobenzene	8.224	168	177625	50.000	ug/l	0.00
34) 1,4-Difluorobenzene	9.100	114	350152	50.000	ug/l	0.00
63) Chlorobenzene-d5	11.865	117	327005	50.000	ug/l	0.00
72) 1,4-Dichlorobenzene-d4	13.788	152	117280	50.000	ug/l	0.00
System Monitoring Compounds						
33) 1,2-Dichloroethane-d4	8.577	65	138317	56.867	ug/l	0.00
Spiked Amount	50.000	Range 74 - 125	Recovery	= 113.740%		
35) Dibromofluoromethane	8.165	113	125041	51.162	ug/l	0.00
Spiked Amount	50.000	Range 75 - 124	Recovery	= 102.320%		
50) Toluene-d8	10.565	98	444417	50.103	ug/l	0.00
Spiked Amount	50.000	Range 86 - 113	Recovery	= 100.200%		
62) 4-Bromofluorobenzene	12.847	95	132122	41.767	ug/l	0.00
Spiked Amount	50.000	Range 77 - 121	Recovery	= 83.540%		
Target Compounds						
						Qvalue
16) Acetone	4.430	43	107361	144.479	ug/l	100
20) Methylene Chloride	5.277	84	96859	44.557	ug/l	93
25) 2-Butanone	7.488	43	12161	11.342	ug/l	99

(#) = qualifier out of range (m) = manual integration (+) = signals summed

Data Path : Z:\voasrv\HPCHEM1\MSVOA_N\Data\VN033125\
Data File : VN086187.D
Acq On : 31 Mar 2025 22:17
Operator : JC\MD
Sample : Q1679-08
Misc : 5.0mL/MSVOA_N/WATER
ALS Vial : 22 Sample Multiplier: 1

Instrument :
MSVOA_N
ClientSampleId :
TP-9

Quant Time: Apr 01 01:41:42 2025
Quant Method : Z:\voasrv\HPCHEM1\MSVOA_N\methods\82N031825W.M
Quant Title : SW846 8260
QLast Update : Wed Mar 19 03:20:56 2025
Response via : Initial Calibration

